

LB103362


ISM02.4

INSTRUMENT ID : CU1

Linear

 μ Abs.:

4340E



A= 0.0000e+000

B= 2.2967e-004

C= 7.3072e-003

Rho= 0.9998244

Accept=Accepted

slope
y intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.014	0.014	29	0.000	29					1
0.05	0.050										
0.20	0.200	0.226	0.026	953	0.0 %	953					13
2.50	2.500	2.496	-0.004	10837	0.0 %	10837					0
5.00	5.000	4.877	-0.123	21203	0.0 %	21203					-2
7.50	7.500	7.610	0.110	33103	0.0 %	33103					1
10.0	10.000	9.976	-0.024	43406	0.0 %	43406					0

RS 06.12.19

LB103362
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	29	-		0 ISM02.4	PPB	6/12/2019 16:01	Std
	0.2 S0.2	953	-		0.2 ISM02.4	PPB	6/12/2019 16:03	Std
	2.5 S2.5	10837	-		2.5 ISM02.4	PPB	6/12/2019 16:06	Std
	5 S5.0	21203	-		5 ISM02.4	PPB	6/12/2019 16:08	Std
	7.5 S7.5	33103	-		7.5 ISM02.4	PPB	6/12/2019 16:10	Std
	10 S10.0	43406	-		10 ISM02.4	PPB	6/12/2019 16:16	Std
ICV	ICV	17241	3.9671 -		ISM02.4	PPB	6/12/2019 16:20	SMPL
ICB	ICB	-99	-0.0154 -		ISM02.4	PPB	6/12/2019 16:22	SMPL
CCV043	CCV043	21235	4.8844 -		ISM02.4	PPB	6/12/2019 16:24	SMPL
CCB043	CCB043	-106	-0.017 -		ISM02.4	PPB	6/12/2019 16:27	SMPL
PB120553BL	PBW553	-40	-0.0019 -		ISM02.4	PPB	6/12/2019 16:29	SMPL
K3147-01	MH0BF8	-26	0.0013 -		ISM02.4	PPB	6/12/2019 16:31	SMPL
K3147-02	MH0BG0	-5	0.0062 -		ISM02.4	PPB	6/12/2019 16:33	SMPL
K3147-03	MH0BG2	-7	0.0057 -		ISM02.4	PPB	6/12/2019 16:36	SMPL
K3147-04	MH0BG4	-26	0.0013 -		ISM02.4	PPB	6/12/2019 16:38	SMPL
K3147-05	MH0BG6	-20	0.0027 -		ISM02.4	PPB	6/12/2019 16:40	SMPL
K3147-06	MH0BG8	-43	-0.0026 -		ISM02.4	PPB	6/12/2019 16:42	SMPL
K3147-07	MH0BH0	-33	-0.0003 -		ISM02.4	PPB	6/12/2019 16:45	SMPL
K3147-08	MH0BH2	-53	-0.0049 -		ISM02.4	PPB	6/12/2019 16:47	SMPL
K3147-09	MH0BH4	-62	-0.0069 -		ISM02.4	PPB	6/12/2019 16:49	SMPL
K3147-10	MH0BH6	-90	-0.0134 -		ISM02.4	PPB	6/12/2019 16:52	SMPL
K3147-11	MH0BH8	-66	-0.0079 -		ISM02.4	PPB	6/12/2019 16:54	SMPL
K3147-12	MH0BJ0	-66	-0.0079 -		ISM02.4	PPB	6/12/2019 16:56	SMPL
K3147-13	MH0BJ2	-87	-0.0127 -		ISM02.4	PPB	6/12/2019 16:58	SMPL
K3147-14	MH0BJ4	-71	-0.009 -		ISM02.4	PPB	6/12/2019 17:01	SMPL
K3147-15	MH0BJ6	-96	-0.0147 -		ISM02.4	PPB	6/12/2019 17:03	SMPL
K3147-16	MH0BJ8	-83	-0.0118 -		ISM02.4	PPB	6/12/2019 17:05	SMPL
K3147-17	MH0BK0	-68	-0.0083 -		ISM02.4	PPB	6/12/2019 17:07	SMPL
K3147-18	MH0BK2	-85	-0.0122 -		ISM02.4	PPB	6/12/2019 17:10	SMPL
K3147-19	MH0BK4	-78	-0.0106 -		ISM02.4	PPB	6/12/2019 17:12	SMPL
CCV044	CCV044	21407	4.9239 -		ISM02.4	PPB	6/12/2019 17:14	SMPL
CCB044	CCB044	-74	-0.0097 -		ISM02.4	PPB	6/12/2019 17:16	SMPL
K3147-20	MH0BK6	-68	-0.0083 -		ISM02.4	PPB	6/12/2019 17:19	SMPL
K3147-21	MH0BK6D	-76	-0.0101 -		ISM02.4	PPB	6/12/2019 17:21	SMPL
K3147-22	MH0BK6S	5096	1.1777 -		ISM02.4	PPB	6/12/2019 17:23	SMPL
PB120554BL	PBW554	-10	0.005 -		ISM02.4	PPB	6/12/2019 17:26	SMPL
K3148-01	MH0BF9	-40	-0.0019 -		ISM02.4	PPB	6/12/2019 17:28	SMPL
K3148-02	MH0BG1	-75	-0.0099 -		ISM02.4	PPB	6/12/2019 17:30	SMPL
K3148-03	MH0BG3	-37	-0.0012 -		ISM02.4	PPB	6/12/2019 17:32	SMPL
K3148-04	MH0BG5	-18	0.0032 -		ISM02.4	PPB	6/12/2019 17:35	SMPL
K3148-05	MH0BG7	-79	-0.0108 -		ISM02.4	PPB	6/12/2019 17:37	SMPL
K3148-06	MH0BG9	-42	-0.0023 -		ISM02.4	PPB	6/12/2019 17:39	SMPL
K3148-07	MH0BH1	-37	-0.0012 -		ISM02.4	PPB	6/12/2019 17:41	SMPL
K3148-08	MH0BH3	-1	0.0071 -		ISM02.4	PPB	6/12/2019 17:44	SMPL
K3148-09	MH0BH5	-84	-0.012 -		ISM02.4	PPB	6/12/2019 17:46	SMPL
K3148-10	MH0BH7	-51	-0.0044 -		ISM02.4	PPB	6/12/2019 17:48	SMPL

INSTRUMENT ID : CV1

K3148-11	MH0BH9	-74	-0.0097 -	ISM02.4	PPB	6/12/2019 17:51	SMPL
K3148-12	MH0BJ1	-73	-0.0095 -	ISM02.4	PPB	6/12/2019 17:53	SMPL
K3148-13	MH0BJ3	-42	-0.0023 -	ISM02.4	PPB	6/12/2019 17:55	SMPL
K3148-14	MH0BJ5	-49	-0.0039 -	ISM02.4	PPB	6/12/2019 17:57	SMPL
K3148-15	MH0BJ7	-81	-0.0113 -	ISM02.4	PPB	6/12/2019 18:00	SMPL
K3148-16	MH0BJ9	-74	-0.0097 -	ISM02.4	PPB	6/12/2019 18:02	SMPL
CCV045	CCV045	21302	4.8998 -	ISM02.4	PPB	6/12/2019 18:04	SMPL
CCB045	CCB045	-75	-0.0099 -	ISM02.4	PPB	6/12/2019 18:06	SMPL
K3148-17	MH0BK1	-62	-0.0069 -	ISM02.4	PPB	6/12/2019 18:09	SMPL
K3148-18	MH0BK3	-75	-0.0099 -	ISM02.4	PPB	6/12/2019 18:11	SMPL
K3148-19	MH0BK5	-64	-0.0074 -	ISM02.4	PPB	6/12/2019 18:13	SMPL
K3148-20	MH0BK7	-57	-0.0058 -	ISM02.4	PPB	6/12/2019 18:15	SMPL
K3148-21	MH0BK7D	-75	-0.0099 -	ISM02.4	PPB	6/12/2019 18:18	SMPL
K3148-22	MH0BK7S	5173	1.1954 -	ISM02.4	PPB	6/12/2019 18:20	SMPL
CCV046	CCV046	21815	5.0176 -	ISM02.4	PPB	6/12/2019 18:22	SMPL
CCB046	CCB046	-57	-0.0058 -	ISM02.4	PPB	6/12/2019 18:25	SMPL