

LB103370


7470A

INSTRUMENT ID: CV1

Linear ▼

 μ Abs.:

43272



A= 0.0000e+000

B= 2.2895e-004

C= -2.5232e-002

Rho= 0.9997724

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.030	-0.030	-22	0.000	-22				
0.2	0.200	0.188	-0.012	931	0.0 %	931				
2.5	2.500	2.511	0.011	11079	0.0 %	11079				
5.0	5.000	5.000	0.000	21950	0.0 %	21950				
7.5	7.500	7.649	0.149	33517	0.0 %	33517				
10.0	10.000	9.882	-0.118	43272	0.0 %	43272				

LB103370 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Jun 2019 09:46:14

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	-22	-	0.0000	7470A	PPB	13 Jun 2019 09:46:28	S	Std
0.2 - 1	50-2	931	-	0.2000	7470A	PPB	13 Jun 2019 09:48:38	S	Std
2.5 - 1	52-5	11079	-	2.5000	7470A	PPB	13 Jun 2019 09:50:46	S	Std
5.0 - 1	55-0	21950	-	5.0000	7470A	PPB	13 Jun 2019 09:52:54	S	Std
7.5 - 1	57-5	33517	-	7.5000	7470A	PPB	13 Jun 2019 09:55:02	S	Std
10.0 - 1	510-0	43272	-	10.0000	7470A	PPB	13 Jun 2019 10:02:47	S	Std
ICV90 - 1	ICV90	18216	4.1454	-	7470A	PPB	13 Jun 2019 10:14:16	U	SMPL
ICB90 - 1	ICB90	-73	-0.0419	-	7470A	PPB	13 Jun 2019 10:16:23	U	SMPL
CCV47 - 1	CCV47	22635	5.1571	-	7470A	PPB	13 Jun 2019 10:18:29	U	SMPL
CCB47 - 1	CCB47	-78	-0.0431	-	7470A	PPB	13 Jun 2019 10:20:35	U	SMPL
CRA - 1	CRA	904	0.1817	-	7470A	PPB	13 Jun 2019 10:22:41	U	SMPL
HighStd - 1	HighStd	43395	9.9102	-	7470A	PPB	13 Jun 2019 10:24:48	U	SMPL
ChkStd - 1	ChkStd	31917	7.2823	-	7470A	PPB	13 Jun 2019 10:26:55	U	SMPL
PB120566BL - 1	PBW	-78	-0.0431	-	7470A	PPB	13 Jun 2019 10:29:03	U	SMPL
PB120566BS - 1	LCSW	17517	3.9854	-	7470A	PPB	13 Jun 2019 10:31:10	U	SMPL
K3159-02 - 1	OR-03-061119-A	-21	-0.0300	-	7470A	PPB	13 Jun 2019 10:33:17	U	SMPL
K3159-02DUP - 1	OR-03-061119-ADUP	173	0.0144	-	7470A	PPB	13 Jun 2019 10:35:25	U	SMPL
K3159-02MS - 1	OR-03-061119-AMS	17304	3.9366	-	7470A	PPB	13 Jun 2019 10:37:33	U	SMPL
K3159-02MSD - 1	OR-03-061119-AMSD	17336	3.9439	-	7470A	PPB	13 Jun 2019 10:39:41	U	SMPL
K3159-04 - 1	OR-03-061119-B	-78	-0.0431	-	7470A	PPB	13 Jun 2019 10:41:48	U	SMPL
CCV48 - 1	CCV48	22988	5.2380	-	7470A	PPB	13 Jun 2019 10:43:55	U	SMPL
CCB48 - 1	CCB48	-101	-0.0484	-	7470A	PPB	13 Jun 2019 10:46:02	U	SMPL
K3265-09 - 1	B-01-(0-37)COMP	106	-0.0010	-	7470A	PPB	13 Jun 2019 10:48:08	U	SMPL
K3265-10 - 1	B-04-COMP	44	-0.0152	-	7470A	PPB	13 Jun 2019 10:50:15	U	SMPL
K3265-11 - 1	B-03-COMP	63	-0.0108	-	7470A	PPB	13 Jun 2019 10:52:22	U	SMPL
K3265-12 - 1	B-02-COMP	86	-0.0055	-	7470A	PPB	13 Jun 2019 10:54:29	U	SMPL
K3281-02 - 1	CONCRETE-COMP	-37	-0.0337	-	7470A	PPB	13 Jun 2019 10:56:36	U	SMPL
K3282-02 - 1	25082-2022	82	-0.0065	-	7470A	PPB	13 Jun 2019 10:58:43	U	SMPL
K3283-03 - 1	200030-200034	109	-0.0003	-	7470A	PPB	13 Jun 2019 11:00:51	U	SMPL
K3285-03 - 1	M-1-OW(0-10)	58	-0.0120	-	7470A	PPB	13 Jun 2019 11:02:59	U	SMPL
PB120567BL - 1	PBW	49	-0.0140	-	7470A	PPB	13 Jun 2019 11:05:08	U	SMPL
PB120567BS - 1	LCSW	17643	4.0142	-	7470A	PPB	13 Jun 2019 11:07:15	U	SMPL
CCV49 - 1	CCV49	22843	5.2048	-	7470A	PPB	13 Jun 2019 11:09:21	U	SMPL
CCB49 - 1	CCB49	-95	-0.0470	-	7470A	PPB	13 Jun 2019 11:11:28	U	SMPL
K3284-02 - 1	BPS1-TT-MW308I-20190609	-104	-0.0490	-	7470A	PPB	13 Jun 2019 11:13:34	U	SMPL
K3284-03 - 1	BPS1-TT-MW310S-20190609	-50	-0.0367	-	7470A	PPB	13 Jun 2019 11:15:41	U	SMPL
K3284-04 - 1	BPS1-TT-MW311S-20190609	-29	-0.0319	-	7470A	PPB	13 Jun 2019 11:17:48	U	SMPL
K3284-05 - 1	BP-TT-AOC22-MW10-20190609	-36	-0.0335	-	7470A	PPB	13 Jun 2019 11:19:55	U	SMPL
K3284-05DUP - 1	BP-TT-AOC22-MW10-20190609DUP	-28	-0.0316	-	7470A	PPB	13 Jun 2019 11:22:02	U	SMPL
K3284-06 - 1	K3284-05MS	17260	3.9265	-	7470A	PPB	13 Jun 2019 11:24:09	U	SMPL
K3284-07 - 1	K3284-05MSD	17700	4.0273	-	7470A	PPB	13 Jun 2019 11:26:17	U	SMPL
K3284-08 - 1	BP-TT-AOC22-MW10-20190609-F	-106	-0.0495	-	7470A	PPB	13 Jun 2019 11:28:25	U	SMPL
K3284-09 - 1	BPS1-TT-MW306D-20190609	-59	-0.0387	-	7470A	PPB	13 Jun 2019 11:30:33	U	SMPL
K3284-10 - 1	BP-TT-AOC22-MW03-20190609	166	0.0128	-	7470A	PPB	13 Jun 2019 11:32:40	U	SMPL
CCV50 - 1	CCV50	23084	5.2599	-	7470A	PPB	13 Jun 2019 11:34:49	U	SMPL
CCB50 - 1	CCB50	-96	-0.0472	-	7470A	PPB	13 Jun 2019 11:36:56	U	SMPL
K3284-11 - 1	BPS1-TT-MW308S-20190609	-37	-0.0337	-	7470A	PPB	13 Jun 2019 11:39:03	U	SMPL
K3284-12 - 1	BP-DUP06-20190609	-41	-0.0346	-	7470A	PPB	13 Jun 2019 11:41:09	U	SMPL
K3284-13 - 1	BP-DUP06-20190609-F	-5	-0.0264	-	7470A	PPB	13 Jun 2019 11:43:16	U	SMPL
K3284-14 - 1	BPS1-TT-MW306I-20190609	-30	-0.0321	-	7470A	PPB	13 Jun 2019 11:45:22	U	SMPL
K3284-15 - 1	BPS1-TT-MW303I2-20190610	-50	-0.0367	-	7470A	PPB	13 Jun 2019 11:47:30	U	SMPL
K3284-16 - 1	BPS1-TT-MW303D-20190610	-85	-0.0447	-	7470A	PPB	13 Jun 2019 11:49:37	U	SMPL
K3284-17 - 1	BP-TT-MW202S-20190610	-59	-0.0387	-	7470A	PPB	13 Jun 2019 11:51:45	U	SMPL
K3284-18 - 1	BP-TT-MW202I-20190610	-72	-0.0417	-	7470A	PPB	13 Jun 2019 11:53:53	U	SMPL
K3284-19 - 1	BP-MH-SW4001-SOUTH-20190610	-75	-0.0424	-	7470A	PPB	13 Jun 2019 11:56:02	U	SMPL
K3284-20 - 1	BP-TT-SW4002-20190610	-51	-0.0369	-	7470A	PPB	13 Jun 2019 11:58:09	U	SMPL
CCV51 - 1	CCV51	23075	5.2579	-	7470A	PPB	13 Jun 2019 12:00:16	U	SMPL
CCB51 - 1	CCB51	-81	-0.0438	-	7470A	PPB	13 Jun 2019 12:02:23	U	SMPL
K3284-21 - 1	BP-TT-SW4003-20190610	-34	-0.0330	-	7470A	PPB	13 Jun 2019 12:04:30	U	SMPL
K3284-22 - 1	BP-TT-SW4004-20190610	-11	-0.0278	-	7470A	PPB	13 Jun 2019 12:06:37	U	SMPL
PB120568BL - 1	PBW	21	-0.0204	-	7470A	PPB	13 Jun 2019 12:08:44	U	SMPL
PB120568BS - 1	LCSW	17484	3.9778	-	7470A	PPB	13 Jun 2019 12:10:51	U	SMPL
K3279-02 - 1	BPS1-TT-MW312S-20190607	-13	-0.0282	-	7470A	PPB	13 Jun 2019 12:12:58	U	SMPL
K3279-03 - 1	BPS1-TT-MW312I-20190607	-44	-0.0353	-	7470A	PPB	13 Jun 2019 12:15:05	U	SMPL
K3279-04 - 1	BPS1-TT-MW302I2-20190607	-66	-0.0403	-	7470A	PPB	13 Jun 2019 12:17:13	U	SMPL
K3279-05 - 1	BPS1-TT-MW307S-20190607	-90	-0.0458	-	7470A	PPB	13 Jun 2019 12:19:21	U	SMPL
K3279-06 - 1	BPS1-TT-MW302D-20190607	-91	-0.0461	-	7470A	PPB	13 Jun 2019 12:21:29	U	SMPL
K3279-07 - 1	BP-TT-MW203I-20190607	-86	-0.0449	-	7470A	PPB	13 Jun 2019 12:23:36	U	SMPL
CCV52 - 1	CCV52	23438	5.3410	-	7470A	PPB	13 Jun 2019 12:25:43	U	SMPL
CCB52 - 1	CCB52	-76	-0.0426	-	7470A	PPB	13 Jun 2019 12:27:51	U	SMPL
K3279-08 - 1	BPS1-TT-MW303I1-20190607	568	0.1048	-	7470A	PPB	13 Jun 2019 12:29:59	U	SMPL
K3279-09 - 1	BP-EB03-20190607	-54	-0.0376	-	7470A	PPB	13 Jun 2019 12:32:08	U	SMPL
K3279-10 - 1	BPS1-TT-MW303I1-20190607-F	90	-0.0046	-	7470A	PPB	13 Jun 2019 12:34:15	U	SMPL
K3280-02 - 1	BP-FW-MW01-20190608	-2	-0.0257	-	7470A	PPB	13 Jun 2019 12:36:22	U	SMPL
K3280-03 - 1	BP-FW-MW02-20190608	-44	-0.0353	-	7470A	PPB	13 Jun 2019 12:38:30	U	SMPL
K3280-04 - 1	BP-FW-MW03-20190608	-69	-0.0410	-	7470A	PPB	13 Jun 2019 12:40:37	U	SMPL
K3280-05 - 1	BP-HN-MW27I-20190608	-74	-0.0422	-	7470A	PPB	13 Jun 2019 12:42:45	U	SMPL
K3280-06 - 1	BP-HN-MW29IR-20190608	-59	-0.0387	-	7470A	PPB	13 Jun 2019 12:44:53	U	SMPL
K3280-07 - 1	BP-HN-MW29D-20190608	-24	-0.0307	-	7470A	PPB	13 Jun 2019 12:47:01	U	SMPL
K3280-08 - 1	BP-MH-SW4001-SOUTH-20190608	-28	-0.0316	-	7470A	PPB	13 Jun 2019 12:49:08	U	SMPL
CCV53 - 1	CCV53	23111	5.2661	-	7470A	PPB	13 Jun 2019 12:51:17	U	SMPL
CCB53 - 1	CCB53	-76	-0.0426	-	7470A	PPB	13 Jun 2019 12:53:25	U	SMPL
K3280-08DUP - 1	BP-MH-SW4001-SOUTH-20190608DUP	-24	-0.0307	-	7470A	PPB	13 Jun 2019 12:55:32	U	SMPL
K3280-09 - 1	K3280-08MS	17679	4.0224	-	7470A	PPB	13 Jun 2019 12:57:39	U	SMPL
K3280-10 - 1	K3280-08MSD	17825	4.0559	-	7470A	PPB	13 Jun 2019 12:59:46	U	SMPL
K3280-11 - 1	BP-TT-SW4002-20190608	-140	-0.0573	-	7470A	PPB	13 Jun 2019 13:01:54	U	SMPL
K3280-12 - 1	BP-TT-SW4003-20190608	-85	-0.0447	-	7470A	PPB	13 Jun 2019 13:04:01	U	SMPL
K3280-13 - 1	BP-FW-MW01-20190608-F	-79	-0.0433	-	7470A	PPB	13 Jun 2019 13:06:08	U	SMPL
PB120566TB - 1	PB120494TB	36	-0.0170	-	7470A	PPB	13 Jun 2019 13:08:15	U	SMPL
K3159-02LX5 - 1		-52	-0.0371	-	7470A	PPB	13 Jun 2019 13:10:23	U	SMPL
K3159-02A - 1		17291	3.9336	-	7470A	PPB	13 Jun 2019 13:12:31	U	SMPL
K3284-05LX5 - 1		-108	-0.0500	-	7470A	PPB	13 Jun 2019 13:14:39	U	SMPL
CCV54 - 1	CCV54	23138	5.2723	-	7470A	PPB	13 Jun 2019 13:20:21	U	SMPL

LB103370 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Jun 2019 09:46:14

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB54 - 1	CCB54	-92	-0.0463		-7470A	PPB	13 Jun 2019 13:22:28	U	SMPL
K3284-05A - 1		17298	3.9352		-7470A	PPB	13 Jun 2019 13:24:36	U	SMPL
K3280-08LX5 - 1		-111	-0.0506		-7470A	PPB	13 Jun 2019 13:26:43	U	SMPL
K3280-08A - 1		17830	4.0570		-7470A	PPB	13 Jun 2019 13:28:51	U	SMPL
CCV55 - 1	CCV55	21313	4.8545		-7470A	PPB	13 Jun 2019 13:30:59	U	SMPL
CCB55 - 1	CCB55	-66	-0.0403		-7470A	PPB	13 Jun 2019 13:33:06	U	SMPL