


LB103378

ISM02.4

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 2.1780e-004

C= -7.2103e-003

Rho= 0.9996261

Accept=Accepted

slope
y intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.035	-0.035	-126	0.000	-126				
0.05	0.050									
0.20	0.200	0.181	-0.019	865	0.0 %	865				
2.50	2.500	2.443	-0.057	11252	0.0 %	11252				
5.00	5.000	5.148	0.148	23668	0.0 %	23668				
7.50	7.500	7.611	0.111	34980	0.0 %	34980				
10.0	10.000	9.851	-0.149	45264	0.0 %	45264				

2.1

-10

-2

3

1

-1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S.0.0	-126	-		0 ISM02.4	PPB	6/13/2019 15:09	Std
	0.2 S0.2	865	-		0.2 ISM02.4	PPB	6/13/2019 15:12	Std
	2.5 S2.5	11252	-		2.5 ISM02.4	PPB	6/13/2019 15:14	Std
	5 S5.0	23668	-		5 ISM02.4	PPB	6/13/2019 15:16	Std
	7.5 S7.5	34980	-		7.5 ISM02.4	PPB	6/13/2019 15:19	Std
	10 S10.0	45264	-		10 ISM02.4	PPB	6/13/2019 15:21	Std
ICV	ICV	18650	4.0547 -		ISM02.4	PPB	6/13/2019 15:25	SMPL
ICB	ICB	-130	-0.0355 -		ISM02.4	PPB	6/13/2019 15:27	SMPL
CCV056	CCV056	23757	5.167 -		ISM02.4	PPB	6/13/2019 15:29	SMPL
CCB056	CCB056	-100	-0.029 -		ISM02.4	PPB	6/13/2019 15:32	SMPL
PB120578BL	PBS578	-49	-0.0179 -		ISM02.4	PPB	6/13/2019 15:34	SMPL
K3206-01	MH0AZ5	7546	1.6363 -		ISM02.4	PPB	6/13/2019 15:36	SMPL
K3206-02	MH0AZ6	48664	10.5916 -		ISM02.4	PPB	6/13/2019 15:38	SMPL
K3206-03	MH0AZ8	454789	HIGH -		ISM02.4	PPB	6/13/2019 15:41	
K3206-04	MH0BA0	50674	11.0294 -		ISM02.4	PPB	6/13/2019 15:42	SMPL
K3206-05	MH0BA1	16689	3.6276 -		ISM02.4	PPB	6/13/2019 15:45	SMPL
K3206-06	MH0BA4	383	0.0762 -		ISM02.4	PPB	6/13/2019 15:47	SMPL
K3206-07	MH0BA5	1659	0.3541 -		ISM02.4	PPB	6/13/2019 15:49	SMPL
K3206-08	MH0BA6	512	0.1043 -		ISM02.4	PPB	6/13/2019 15:52	SMPL
K3206-09	MH0BA7	512	0.1043 -		ISM02.4	PPB	6/13/2019 15:54	SMPL
K3206-10	MH0BA8	1277	0.2709 -		ISM02.4	PPB	6/13/2019 15:56	SMPL
K3206-11	MH0BA9	4798	1.0378 -		ISM02.4	PPB	6/13/2019 15:58	SMPL
K3206-12	MH0BB0	1970	0.4218 -		ISM02.4	PPB	6/13/2019 16:01	SMPL
K3206-13	MH0BB1	6387	1.3839 -		ISM02.4	PPB	6/13/2019 16:03	SMPL
K3206-14	MH0BB2	4473	0.967 -		ISM02.4	PPB	6/13/2019 16:05	SMPL
K3206-15	MH0BB3	24782	5.3902 -		ISM02.4	PPB	6/13/2019 16:07	SMPL
K3206-16	MH0BB4	2768	0.5957 -		ISM02.4	PPB	6/13/2019 16:10	SMPL
K3206-17	MH0BB6	244	0.0459 -		ISM02.4	PPB	6/13/2019 16:12	SMPL
K3206-18	MH0BB7	1122	0.2372 -		ISM02.4	PPB	6/13/2019 16:14	SMPL
K3206-19	MH0BB8	2295	0.4926 -		ISM02.4	PPB	6/13/2019 16:17	SMPL
CCV057	CCV057	23902	5.1986 -		ISM02.4	PPB	6/13/2019 16:19	SMPL
CCB057	CCB057	-101	-0.0292 -		ISM02.4	PPB	6/13/2019 16:21	SMPL
K3206-20	MH0BB9	6541	1.4174 -		ISM02.4	PPB	6/13/2019 16:23	SMPL
K3206-21	MH0BB9D	4925	1.0654 -		ISM02.4	PPB	6/13/2019 16:26	SMPL
K3206-22	MH0BB9S	18174	3.951 -		ISM02.4	PPB	6/13/2019 16:28	SMPL
PB120579BL	PBS579	-98	-0.0286 -		ISM02.4	PPB	6/13/2019 16:30	SMPL
K3207-01	MH0BC0	4507	0.9744 -		ISM02.4	PPB	6/13/2019 16:32	SMPL
K3207-02	MH0BC1	11335	2.4615 -		ISM02.4	PPB	6/13/2019 16:35	SMPL
K3207-03	MH0BC2	5668	1.2273 -		ISM02.4	PPB	6/13/2019 16:37	SMPL
K3207-04	MH0BC3	10022	2.1755 -		ISM02.4	PPB	6/13/2019 16:39	SMPL
K3207-05	MH0BC5	496	0.1008 -		ISM02.4	PPB	6/13/2019 16:42	SMPL
K3207-06	MH0BC6	5665	1.2266 -		ISM02.4	PPB	6/13/2019 16:44	SMPL
K3207-07	MH0BE8	2469	0.5305 -		ISM02.4	PPB	6/13/2019 16:46	SMPL
K3207-08	MH0BE9	9874	2.1433 -		ISM02.4	PPB	6/13/2019 16:48	SMPL
K3207-09	MH0BF0	10635	2.3091 -		ISM02.4	PPB	6/13/2019 16:51	SMPL
K3207-10	MH0BF1	1154	0.2441 -		ISM02.4	PPB	6/13/2019 16:53	SMPL

INSTRUMENT ID : CV1

K3207-11	MH0BF2	1537	0.3275 -	ISM02.4	PPB	6/13/2019 16:55	SMPL
K3207-12	MH0BF4	16231	3.5278 -	ISM02.4	PPB	6/13/2019 16:57	SMPL
K3207-13	MH0BC7	667	0.1381 -	ISM02.4	PPB	6/13/2019 17:00	SMPL
K3207-14	MH0BC8	4873	1.0541 -	ISM02.4	PPB	6/13/2019 17:02	SMPL
K3207-15	MH0BC9	1915	0.4099 -	ISM02.4	PPB	6/13/2019 17:04	SMPL
K3207-16	MH0BD0	860	0.1801 -	ISM02.4	PPB	6/13/2019 17:06	SMPL
CCV058	CCV058	22851	4.9697 -	ISM02.4	PPB	6/13/2019 17:09	SMPL
CCB058	CCB058	-101	-0.0292 -	ISM02.4	PPB	6/13/2019 17:11	SMPL
K3207-17	MH0BD1	5576	1.2072 -	ISM02.4	PPB	6/13/2019 17:13	SMPL
K3207-18	MH0BD2	1777	0.3798 -	ISM02.4	PPB	6/13/2019 17:16	SMPL
K3207-19	MH0BD3	1500	0.3195 -	ISM02.4	PPB	6/13/2019 17:18	SMPL
K3207-20	MH0BD4	5714	1.2373 -	ISM02.4	PPB	6/13/2019 17:20	SMPL
K3207-21	MH0BD4D	5666	1.2268 -	ISM02.4	PPB	6/13/2019 17:22	SMPL
K3207-22	MH0BD4S	15116	3.285 -	ISM02.4	PPB	6/13/2019 17:28	SMPL
K3205-15	MH0AY8	2240	0.4807 -	ISM02.4	PPB	6/13/2019 17:31	SMPL
K3205-16	MH0AY9	3631	0.7836 -	ISM02.4	PPB	6/13/2019 17:33	SMPL
K3205-18	MH0AZ2	8242	1.7879 -	ISM02.4	PPB	6/13/2019 17:35	SMPL
K3205-19	MH0AZ3	10878	2.362 -	ISM02.4	PPB	6/13/2019 17:37	SMPL
K3206-02DLX5		9151	1.9858 -	ISM02.4	PPB	6/13/2019 17:40	SMPL
K3206-03DLX10		69841	15.2039 -	ISM02.4	PPB	6/13/2019 17:42	SMPL
K3206-04DLX5		9468	2.0549 -	ISM02.4	PPB	6/13/2019 17:44	SMPL
k3206-03DLX50		13581	2.9507 -	ISM02.4	PPB	6/13/2019 17:50	SMPL
CCV059	CCV059	23418	5.0931 -	ISM02.4	PPB	6/13/2019 17:52	SMPL
CCB059	CCB059	-89	-0.0266 -	ISM02.4	PPB	6/13/2019 17:54	SMPL