



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103390

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/13/2019	17:35
2	CAL2	1	Water	NA	NA	20.2	7.00	06/13/2019	17:36
3	CAL3	1	Water	NA	NA	20.2	10.04	06/13/2019	17:38
4	ICV	1	Water	NA	NA	20.1	7.01	06/13/2019	17:40
5	CCV1	1	Water	NA	NA	20.2	2.01	06/13/2019	18:15
6	K3369-02	1	Water	NA	NA	20.1	6.97	06/13/2019	18:16
7	K3369-02DUP	1	Water	NA	NA	20.2	6.98	06/13/2019	18:17
8	CCV2	1	Water	NA	NA	20.2	12.02	06/13/2019	18:21

29 103396

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k3373

WorkList ID : 127396

Date : 06-13-2019 17:37:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/13/2019	Water	K3369-02	Corrosivity	Cool 4 deg C	PSEG01	O31	FRAC-TANK-COMP	06/13/2019	9040C

Date/Time 06.13.19 17:40
 Received by: Jn
 Relinquished by: pc

Date/Time 06.13.19 18:25
 Received by:
 Relinquished by: Jn