



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103391

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/13/2019	17:35
2	CAL2	1	Water	NA	NA	20.2	7.00	06/13/2019	17:36
3	CAL3	1	Water	NA	NA	20.2	10.04	06/13/2019	17:38
4	ICV	1	Water	NA	NA	20.1	7.01	06/13/2019	17:40
5	CCV1	1	Water	NA	NA	20.2	2.01	06/13/2019	18:25
6	K3373-05	1	Solid	20.02	20	22.3	7.64	06/13/2019	18:26
7	K3373-05DUP	1	Solid	20.01	20	22.4	7.65	06/13/2019	18:29
8	CCV2	1	Water	NA	NA	20.2	12.01	06/13/2019	18:30

23103391

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K3373 WorkList ID : 127397 Date : 06-13-2019 17:38:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/13/2019	Solid	K3373-05	pH	Cool 4 deg C	PSEG01	O31	13803-13804	06/13/2019	9041A

Date/Time 06-13-19 17:40
Received by: J
Relinquished by: PC

Date/Time 06-13-19 18:25
Received by: PC
Relinquished by: Jo