


LB103395

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.8100e-004

C= -1.1335e-002

Rho= 0.9998095

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-40	0.000	-40				
0.2	0.200	0.187	-0.013	1097	0.0 %	1097				
2.5	2.500	2.449	-0.051	13595	0.0 %	13595				
5.0	5.000	5.158	0.158	28559	0.0 %	28559				
7.5	7.500	7.473	-0.027	41349	0.0 %	41349				
10.0	10.000	9.951	-0.049	55041	0.0 %	55041				

LB103395 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Jun 2019 18:10:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-40	-	0.00007471B	PPB		13 Jun 2019 18:10:12	S	Std
0.2 - 1	50.2	1097	-	0.20007471B	PPB		13 Jun 2019 18:12:30	S	Std
2.5 - 1	52.5	13595	-	2.50007471B	PPB		13 Jun 2019 18:14:48	S	Std
5.0 - 1	55.0	28559	-	5.00007471B	PPB		13 Jun 2019 18:17:06	S	Std
7.5 - 1	57.5	41349	-	7.50007471B	PPB		13 Jun 2019 18:19:22	S	Std
10.0 - 1	510.0	55041	-	10.00007471B	PPB		13 Jun 2019 18:25:30	S	Std
ICV93 - 1	ICV93	22916	4.1365	-7471B	PPB		13 Jun 2019 18:29:25	U	SMPL
ICB93 - 1	ICB93	-153	-0.0390	-7471B	PPB		13 Jun 2019 18:31:41	U	SMPL
CCV60 - 1	CCV60	29692	5.3630	-7471B	PPB		13 Jun 2019 18:33:56	U	SMPL
CCB60 - 1	CCB60	-150	-0.0385	-7471B	PPB		13 Jun 2019 18:36:12	U	SMPL
CRA - 1	CRA	1082	0.1845	-7471B	PPB		13 Jun 2019 18:38:27	U	SMPL
HighStd - 1	HighStd	55911	10.1086	-7471B	PPB		13 Jun 2019 18:40:42	U	SMPL
ChkStd - 1	ChkStd	39914	7.2132	-7471B	PPB		13 Jun 2019 18:42:58	U	SMPL
PB120593BL - 1	PBS	-195	-0.0466	-7471B	PPB		13 Jun 2019 18:45:16	U	SMPL
PB120593BS - 1	LCSS	20761	3.7464	-7471B	PPB		13 Jun 2019 18:47:33	U	SMPL
K3156-03 - 1	HD-01-061219	1982	0.3474	-7471B	PPB		13 Jun 2019 18:49:50	U	SMPL
K3156-03DUP - 1	HD-01-061219DUP	1957	0.3429	-7471B	PPB		13 Jun 2019 18:52:07	U	SMPL
K3156-03MS - 1	HD-01-061219MS	25166	4.5438	-7471B	PPB		13 Jun 2019 18:54:24	U	SMPL
K3156-03MSD - 1	HD-01-061219MSD	26279	4.7452	-7471B	PPB		13 Jun 2019 18:56:41	U	SMPL
K3294-01 - 1	S1(0-5)	2836	0.5020	-7471B	PPB		13 Jun 2019 18:58:57	U	SMPL
CCV61 - 1	CCV61	26778	4.8355	-7471B	PPB		13 Jun 2019 19:01:12	U	SMPL
CCB61 - 1	CCB61	-137	-0.0361	-7471B	PPB		13 Jun 2019 19:03:32	U	SMPL
K3294-02 - 1	S1(5-10)	498	0.0788	-7471B	PPB		13 Jun 2019 19:05:48	U	SMPL
K3294-03 - 1	S2(0-5)	1007	0.1709	-7471B	PPB		13 Jun 2019 19:08:04	U	SMPL
K3294-04 - 1	S2(5-10)	255	0.0348	-7471B	PPB		13 Jun 2019 19:10:19	U	SMPL
K3333-01 - 1	CONCRETE-1	207	0.0261	-7471B	PPB		13 Jun 2019 19:12:35	U	SMPL
K3333-02 - 1	CONCRETE-2	28	-0.0063	-7471B	PPB		13 Jun 2019 19:14:51	U	SMPL
K3333-03 - 1	CONCRETE-3	50	-0.0023	-7471B	PPB		13 Jun 2019 19:17:08	U	SMPL
K3333-04 - 1	CONCRETE-4	434	0.0672	-7471B	PPB		13 Jun 2019 19:19:25	U	SMPL
K3333-05 - 1	CONCRETE-5	577	0.0931	-7471B	PPB		13 Jun 2019 19:21:42	U	SMPL
K3333-06 - 1	CONCRETE-6	33	-0.0054	-7471B	PPB		13 Jun 2019 19:24:00	U	SMPL
K3342-01 - 1	FK-EF-01-012-ABCD	5726	1.0251	-7471B	PPB		13 Jun 2019 19:26:16	U	SMPL
CCV62 - 1	CCV62	27353	4.9396	-7471B	PPB		13 Jun 2019 19:28:32	U	SMPL
CCB62 - 1	CCB62	-106	-0.0305	-7471B	PPB		13 Jun 2019 19:30:48	U	SMPL
K3343-01 - 1	FK-0D015-03-008	1370	0.2366	-7471B	PPB		13 Jun 2019 19:33:05	U	SMPL
K3344-01 - 1	FK-0D015-03-007	1371	0.2368	-7471B	PPB		13 Jun 2019 19:35:21	U	SMPL
K3346-01 - 1	RICH-REC-ROAD-STONE-15	1961	0.3436	-7471B	PPB		13 Jun 2019 19:37:36	U	SMPL
K3156-03LX5 - 1		197	0.0243	-7471B	PPB		13 Jun 2019 19:39:52	U	SMPL
K3156-03A - 1		25524	4.6086	-7471B	PPB		13 Jun 2019 19:42:09	U	SMPL
CCV63 - 1	CCV63	26093	4.7115	-7471B	PPB		13 Jun 2019 19:48:16	U	SMPL
CCB63 - 1	CCB63	-107	-0.0307	-7471B	PPB		13 Jun 2019 19:50:34	U	SMPL