



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/14/2019

Run Number: LB103398

BalanceID: WC SC-3

OvenID: WC-OVEN-2

FilterID: 14466061

ThermometerID: WET-OVEN 2

TEMP1 IN: 103 06/13/2019 16:00 TEMP OUT: 104 06/14/2019 07:00
TEMP2 IN: 104 06/14/2019 07:45 TEMP OUT: 104 06/14/2019 09:00
TEMP3 IN: 105 06/14/2019 10:10 TEMP OUT: 105 06/14/2019 11:30
TEMP3 IN: 105 06/14/2019 12:20 TEMP OUT: 105 06/14/2019 14:00

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB103398BL	LB103398BL	1.4052	1.4052	100	1.4052	1.4052	1.4052	0	0
2	LB103398BS	LB103398BS	1.3986	1.3986	100	1.449	1.449	1.449	0.0504	504
3	K3336-01	001M-WILLETS-PT-BLVD (JUNE)	1.3869	1.3869	150	1.394	1.394	1.394	0.0071	47.3
4	K3336-02	002M-35TH-AVE (JUNE)	1.3804	1.3804	200	1.3867	1.3867	1.3867	0.0063	31.5
5	K3348-01	TANKA 190612	1.388	1.388	400	1.3925	1.3925	1.3925	0.0045	11.2
6	K3348-01DUP	TANKA_190612DUP	1.3848	1.3848	400	1.3893	1.3893	1.3893	0.0045	11.2
7	K3348-02	TANKB_190612	1.3924	1.3924	400	1.3931	1.3931	1.3931	0.0007	1.8
8	K3348-03	TANKC 190612	1.3908	1.3908	250	1.3925	1.3925	1.3925	0.0017	6.8
9	K3350-03	OUTFALL001-190611	1.3871	1.3871	900	1.3874	1.3874	1.3874	0.0003	0.3
10	K3354-01	061119-JETT-E-D-B	1.3866	1.3866	300	1.4001	1.4001	1.4001	0.0135	45
11	K3354-02	061119-JETT-E-D-M	1.3926	1.3926	500	1.4047	1.4047	1.4047	0.0121	24.2
12	K3354-03	061119-JETT-E-D-S	1.3898	1.3898	500	1.3983	1.3983	1.3983	0.0085	17
13	K3354-04	061119-JETT-E-U-B	1.3864	1.3864	500	1.403	1.403	1.403	0.0166	33.2
14	K3354-05	061119-JETT-E-U-M	1.3958	1.3958	500	1.4071	1.4071	1.4071	0.0113	22.6
15	K3354-06	061119-JETT-E-U-S	1.3815	1.3815	500	1.3926	1.3926	1.3926	0.0111	22.2



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 06/14/2019Run Number: LB103398BalanceID: WC SC-3OvenID: WC-OVEN-2FilterID: 14466061ThermometerID: WET-OVEN 2

TEMP1 IN: 103 06/13/2019 16:00 TEMP OUT: 104 06/14/2019 07:00
TEMP2 IN: 104 06/14/2019 07:45 TEMP OUT: 104 06/14/2019 09:00
TEMP3 IN: 105 06/14/2019 10:10 TEMP OUT: 105 06/14/2019 11:30
TEMP3 IN: 105 06/14/2019 12:20 TEMP OUT: 105 06/14/2019 14:00

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	K3354-07	061119-SIDI-E-D-B	1.3932	1.3932	300	1.4107	1.4107	1.4107	0.0175	58.3
17	K3354-08	061119-SIDI-E-D-M	1.3838	1.3838	300	1.4027	1.4027	1.4027	0.0189	63
18	K3354-09	061119-SIDI-E-D-S	1.3936	1.3936	300	1.4118	1.4118	1.4118	0.0182	60.7
19	K3354-10	061119-SIDI-E-U-B	1.3848	1.3848	100	1.4025	1.4025	1.4025	0.0177	177
20	K3354-11	061119-SIDI-E-U-M	1.3865	1.3865	300	1.4062	1.4062	1.4062	0.0197	65.7
21	K3354-12	061119-SIDI-E-U-S	1.3833	1.3833	300	1.3972	1.3972	1.3972	0.0139	46.3
22	K3354-12DUP	061119-SIDI-E-U-SDUP	1.3959	1.3959	300	1.4099	1.4099	1.4099	0.014	46.7
23	K3369-01	FRAC-TANK-COMP	1.3874	1.3874	500	1.3972	1.3972	1.3972	0.0098	19.6

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS061419

WorkList ID : 127402

Date : 06-14-2019 07:33:05

2/103398

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/17/2019	Water	K3336-01	TSS	Cool 4 deg C	TULL01	H42	001M-WILLETTS-PT-BLVD(JUL)	06/12/2019	SM2540 D
06/17/2019	Water	K3336-02	TSS	Cool 4 deg C	TULL01	H42	002M-35TH-AVE(JUNE)	06/12/2019	SM2540 D
06/19/2019	Water	K3348-01	TSS	Cool 4 deg C	ENVI77	H52	TANKA_190612	06/12/2019	SM2540 D
06/19/2019	Water	K3348-02	TSS	Cool 4 deg C	ENVI77	H52	TANKB_190612	06/12/2019	SM2540 D
06/19/2019	Water	K3348-03	TSS	Cool 4 deg C	ENVI77	H52	TANKC_190612	06/12/2019	SM2540 D
06/18/2019	Water	K3350-03	TSS	Cool 4 deg C	SHAW22	H11	OUTFALL001-190611	06/11/2019	SM2540 D
06/18/2019	Water	K3354-01	TSS	Cool 4 deg C	HDR103	H11	061119-JETT-E-D-B	06/11/2019	SM2540 D
06/18/2019	Water	K3354-02	TSS	Cool 4 deg C	HDR103	H11	061119-JETT-E-D-M	06/11/2019	SM2540 D
06/18/2019	Water	K3354-03	TSS	Cool 4 deg C	HDR103	H11	061119-JETT-E-D-S	06/11/2019	SM2540 D
06/18/2019	Water	K3354-04	TSS	Cool 4 deg C	HDR103	H11	061119-JETT-E-U-B	06/11/2019	SM2540 D
06/18/2019	Water	K3354-05	TSS	Cool 4 deg C	HDR103	H11	061119-JETT-E-U-M	06/11/2019	SM2540 D
06/18/2019	Water	K3354-06	TSS	Cool 4 deg C	HDR103	H11	061119-JETT-E-U-S	06/11/2019	SM2540 D
06/18/2019	Water	K3354-07	TSS	Cool 4 deg C	HDR103	H11	061119-SIDI-E-D-B	06/11/2019	SM2540 D
06/18/2019	Water	K3354-08	TSS	Cool 4 deg C	HDR103	H11	061119-SIDI-E-D-M	06/11/2019	SM2540 D
06/18/2019	Water	K3354-09	TSS	Cool 4 deg C	HDR103	H11	061119-SIDI-E-D-S	06/11/2019	SM2540 D
06/18/2019	Water	K3354-10	TSS	Cool 4 deg C	HDR103	H11	061119-SIDI-E-U-B	06/11/2019	SM2540 D
06/18/2019	Water	K3354-11	TSS	Cool 4 deg C	HDR103	H11	061119-SIDI-E-U-M	06/11/2019	SM2540 D
06/18/2019	Water	K3354-12	TSS	Cool 4 deg C	HDR103	H11	061119-SIDI-E-U-S	06/11/2019	SM2540 D
06/19/2019	Water	K3369-01	TSS	Cool 4 deg C	PSEG01	O31	FRAC-TANK-COMP	06/13/2019	SM2540 D

Date/Time 06/14/19 09:00
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 06/14/19 10:30
 Received by: [Signature]
 Relinquished by: [Signature]