

Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB103399

Reagent/Standard	Lot/Loq #
MBAS CAL STD, 0.5PPM	WP75499
MBAS CCV STD, 1.0PPM	WP75505
MBAS ICV-LCS std 1.0PPM	WP75506
MBAS DAILY CHECK STD, 0.01PPM	WP75507
MBAS DAILY CHECK STD, 2.0PPM	WP75508
MBAS CAL STD 2 PPM	WP75496
MBAS CAL STD 1 PPM	WP75498
MBAS CAL STD 0.1 PPM	WP75502
mbas cal std 0.05 ppm	WP75503
MBAS CAL STD 0.01 PPM	WP75504
MBAS BLANK 0 PPM	WP75495
MBAS CAL STD 0.4 PPM	WP75500
MBAS CAL STD 0.2 PPM	WP75501
MBAS Wash Solution	WP73117
Methylene blue reagent-1000ml	WP73119

Intercept: -0.0001

Slope: 0.221

Regression: 0.9992

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Anal Date	Anal Time
1	CAL1	0	1	0.000	06/14/2019	08:15
2	CAL2	0.01	1	0.002	06/14/2019	08:15
3	CAL3	0.05	1	0.011	06/14/2019	08:16
4	CAL4	0.1	1	0.022	06/14/2019	08:17
5	CAL5	0.2	1	0.039	06/14/2019	08:17
6	CAL6	0.4	1	0.087	06/14/2019	08:18
7	CAL7	0.5	1	0.108	06/14/2019	08:19
8	CAL8	1	1	0.218	06/14/2019	08:19
9	CAL9	2	1	0.436	06/14/2019	08:20



Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB103399

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Result (mg/L)	AnalDate	Anal Time
1	ICV	1	1	0.211	0.955	06/14/2019	08:21
2	ICB		1	0.000	0.000	06/14/2019	08:21
3	CHK1	0.01	1	0.002	0.010	06/14/2019	08:22
4	CHK2	2	1	0.451	2.041	06/14/2019	08:23
5	CCV1	1	1	0.232	1.050	06/14/2019	08:24
6	CCB1		1	0.000	0.000	06/14/2019	08:25
7	LB103399BL		1	0.000	0.000	06/14/2019	08:25
8	LB103399BS	1	1	0.219	0.991	06/14/2019	08:26
9	K3348-01		2	0.274	1.240	06/14/2019	08:27
10	K3348-02		1	0.285	1.290	06/14/2019	08:28
11	K3348-03		1	0.020	0.091	06/14/2019	08:30
12	K3348-03DUP		1	0.021	0.095	06/14/2019	08:31
13	K3348-03MS	0.5	1	0.157	0.711	06/14/2019	08:32
14	K3348-03MSD	0.5	1	0.150	0.679	06/14/2019	08:33
15	CCV2	1	1	0.229	1.037	06/14/2019	08:34
16	CCB2		1	0.000	0.000	06/14/2019	08:35

WORKLIST(Hardcopy Internal Chain)

03103399

WorkList Name : mbas061319

WorkList ID : 127382

Date : 06-13-2019 13:31:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/14/2019	Water	K3348-01	MBAS	Cool 4 deg C	ENVI77	H52	TANKA_190612	06/12/2019	SM5540 C
06/14/2019	Water	K3348-02	MBAS	Cool 4 deg C	ENVI77	H52	TANKB_190612	06/12/2019	SM5540 C
06/14/2019	Water	K3348-03	MBAS	Cool 4 deg C	ENVI77	H52	TANKC_190612	06/12/2019	SM5540 C

Date/Time 06/14/19 7:45
 Received by: 
 Relinquished by: 

Date/Time 06/14/19 9:00
 Received by: 
 Relinquished by: 