

OVEN TEMP1: 103 °C TEMP: 105 °C

DATE & TIME IN: 06/13/2019 15:00 OUT: 06/14/2019 07:10

OVEN TEMP2: 105 °C TEMP: 105 °C

DATE & TIME IN: 06/14/2019 08:30 OUT: 06/14/2019 10:05

Run Number: LB103400

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB103400BL	LB103400BL	90.1524	90.1524	100	90.1524	90.1524	90.1524	0	0
2	K3348-01	TANKA_190612	74.8344	74.8344	100	75.2521	75.2521	75.2521	0.4177	4177
3	K3348-01DUP	TANKA_190612DUP	90.065	90.065	100	90.5015	90.5015	90.5015	0.4365	4365
4	K3348-02	TANKB_190612	95.089	95.089	100	95.1277	95.1277	95.1277	0.0387	387
5	K3348-03	TANKC_190612	90.7989	90.7989	100	90.813	90.813	90.813	0.0141	141

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ts061319

WorkList ID : 127374

Date : 06-13-2019 10:06:48

XB103400

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2019	Water	K3348-01	TS	Cool 4 deg C	ENVI77	H52	TANKA_190612	06/12/2019	SM2540 B
06/19/2019	Water	K3348-02	TS	Cool 4 deg C	ENVI77	H52	TANKB_190612	06/12/2019	SM2540 B
06/19/2019	Water	K3348-03	TS	Cool 4 deg C	ENVI77	H52	TANKC_190612	06/12/2019	SM2540 B

Date/Time

Received by:

Relinquished by:

06/13/19 14:00
 Received by: AS
 Relinquished by: AS

Date/Time

Received by:

Relinquished by:

06/13/19 15:30
 Received by: AS
 Relinquished by: AS