

LB103402


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

61140



A= 0.0000e+000

B= 1.6317e-004

C= -3.1326e-002

Rho= 0.9999449

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.036	-0.036	-29	0.000	-29				
0.2	0.200	0.188	-0.012	1342	0.0 %	1342				
2.5	2.500	2.517	0.017	15617	0.0 %	15617				
5.0	5.000	5.059	0.059	31196	0.0 %	31196				
7.5	7.500	7.528	0.028	46325	0.0 %	46325				
10.0	10.000	9.945	-0.055	61140	0.0 %	61140				

LB103402 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Jun 2019 10:06:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	-29	-	0.0000	7471B	PPB	14 Jun 2019 10:13:48	S	Std
0.2 - 1	50-2	1342	-	0.2000	7471B	PPB	14 Jun 2019 10:16:05	S	Std
2.5 - 1	52-5	15617	-	2.5000	7471B	PPB	14 Jun 2019 10:18:23	S	Std
5.0 - 1	55-0	31196	-	5.0000	7471B	PPB	14 Jun 2019 10:20:40	S	Std
7.5 - 1	57-5	46325	-	7.5000	7471B	PPB	14 Jun 2019 10:22:57	S	Std
10.0 - 1	510-0	61140	-	10.0000	7471B	PPB	14 Jun 2019 10:31:51	S	Std
ICV94 - 1	ICV94	25942	4.2017	-	7471B	PPB	14 Jun 2019 10:35:27	U	SMPL
ICB94 - 1	ICB94	-69	-0.0426	-	7471B	PPB	14 Jun 2019 10:37:45	U	SMPL
CCV64 - 1	CCV64	31822	5.1611	-	7471B	PPB	14 Jun 2019 10:39:59	U	SMPL
CCB64 - 1	CCB64	-64	-0.0418	-	7471B	PPB	14 Jun 2019 10:42:15	U	SMPL
CRA - 1	CRA	1392	0.1958	-	7471B	PPB	14 Jun 2019 10:44:30	U	SMPL
HighStd - 1	HighStd	62275	10.1302	-	7471B	PPB	14 Jun 2019 10:46:46	U	SMPL
ChkStd - 1	ChkStd	43525	7.0707	-	7471B	PPB	14 Jun 2019 10:49:02	U	SMPL
PB120619BL - 1	PBS	-111	-0.0494	-	7471B	PPB	14 Jun 2019 10:51:21	U	SMPL
PB120619BS - 1	LCSS	23379	3.7835	-	7471B	PPB	14 Jun 2019 10:53:39	U	SMPL
K3351-01 - 1	RT-3407	1795	0.2616	-	7471B	PPB	14 Jun 2019 10:55:55	U	SMPL
K3351-01DUP - 1	RT-3407DUP	1586	0.2275	-	7471B	PPB	14 Jun 2019 10:58:12	U	SMPL
K3351-01MS - 1	RT-3407MS	25910	4.1964	-	7471B	PPB	14 Jun 2019 11:00:29	U	SMPL
K3351-01MSD - 1	RT-3407MSD	26188	4.2418	-	7471B	PPB	14 Jun 2019 11:02:48	U	SMPL
K3351-03 - 1	RT-3425	52351	8.5108	-	7471B	PPB	14 Jun 2019 11:05:04	U	SMPL
CCV65 - 1	CCV65	32335	5.2448	-	7471B	PPB	14 Jun 2019 11:07:21	U	SMPL
CCB65 - 1	CCB65	-159	-0.0573	-	7471B	PPB	14 Jun 2019 11:09:38	U	SMPL
K3351-05 - 1	RBR-200029	943	0.1225	-	7471B	PPB	14 Jun 2019 11:11:55	U	SMPL
K3351-07 - 1	RT-3860	1701	0.2462	-	7471B	PPB	14 Jun 2019 11:14:10	U	SMPL
K3361-02 - 1	BAT-B10	722	0.0865	-	7471B	PPB	14 Jun 2019 11:16:26	U	SMPL
K3361-04 - 1	BAT-B15	66721	10.8556	-	7471B	PPB	14 Jun 2019 11:18:42	U	SMPL
K3366-01 - 1	OILY-SOIL	3381	0.5204	-	7471B	PPB	14 Jun 2019 11:20:59	U	SMPL
K3369-04 - 1	Z-3-SOIL-1	3237	0.4969	-	7471B	PPB	14 Jun 2019 11:23:18	U	SMPL
K3369-08 - 1	Z-3-SOIL-2	1197	0.1640	-	7471B	PPB	14 Jun 2019 11:25:35	U	SMPL
K3369-12 - 1	Z-3-SOIL-3	5651	0.8908	-	7471B	PPB	14 Jun 2019 11:27:52	U	SMPL
K3369-16 - 1	Z-3-SOIL-4	274	0.0134	-	7471B	PPB	14 Jun 2019 11:30:10	U	SMPL
K3373-05 - 1	13803-13804	4987	0.7824	-	7471B	PPB	14 Jun 2019 11:32:26	U	SMPL
CCV66 - 1	CCV66	32275	5.2350	-	7471B	PPB	14 Jun 2019 11:34:42	U	SMPL
CCB66 - 1	CCB66	-96	-0.0470	-	7471B	PPB	14 Jun 2019 11:36:58	U	SMPL
K3351-01LX5 - 1		203	0.0018	-	7471B	PPB	14 Jun 2019 11:39:13	U	SMPL
K3351-01A - 1		26196	4.2431	-	7471B	PPB	14 Jun 2019 11:41:29	U	SMPL
K3361-04DLX2 - 1		33115	5.3721	-	7471B	PPB	14 Jun 2019 11:43:45	U	SMPL
CCV67 - 1	CCV67	30032	4.8690	-	7471B	PPB	14 Jun 2019 11:48:55	U	SMPL
CCB67 - 1	CCB67	-103	-0.0481	-	7471B	PPB	14 Jun 2019 11:51:11	U	SMPL