

LB103406


ISM02.4

INSTRUMENT ID : CUG

Linear

 μ Abs.:

52624



A= 0.0000e+000

B= 1.9135e-004

C= -5.5315e-003

Rho= 0.9997445

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.027	0.027	169	0.000	169				
0.05	0.050									
0.20	0.200	0.221	0.021	1186	0.0 %	1186				
2.50	2.500	2.560	0.060	13405	0.0 %	13405				
5.00	5.000	4.819	-0.181	25212	0.0 %	25212				
7.50	7.500	7.509	0.009	39270	0.0 %	39270				
10.0	10.000	10.064	0.064	52624	0.0 %	52624				

%D

—
11
2
-4
0
1

RS 06.14.19

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	169	-		0 ISM02.4	PPB	6/14/2019 12:29	Std
	0.2 S0.2	1186	-		0.2 ISM02.4	PPB	6/14/2019 12:31	Std
	2.5 S2.5	13405	-		2.5 ISM02.4	PPB	6/14/2019 12:33	Std
	5 S5.0	25212	-		5 ISM02.4	PPB	6/14/2019 12:35	Std
	7.5 S7.5	39270	-		7.5 ISM02.4	PPB	6/14/2019 12:38	Std
	10 S10.0	52624	-		10 ISM02.4	PPB	6/14/2019 12:43	Std
ICV	ICV	20424	3.9027 -		ISM02.4	PPB	6/14/2019 12:49	SMPL
ICB	ICB	96	0.0128 -		ISM02.4	PPB	6/14/2019 12:52	SMPL
CCV068	CCV068	25205	4.8176 -		ISM02.4	PPB	6/14/2019 12:54	SMPL
CCB068	CCB068	35	0.0012 -		ISM02.4	PPB	6/14/2019 12:56	SMPL
PB120626BL	PBS626	68	0.0075 -		ISM02.4	PPB	6/14/2019 12:58	SMPL
K3208-01	MH0BD5	7599	1.4486 -		ISM02.4	PPB	6/14/2019 13:01	SMPL
K3208-02	MH0BD6	1106	0.2061 -		ISM02.4	PPB	6/14/2019 13:03	SMPL
K3208-03	MH0BD7	8859	1.6897 -		ISM02.4	PPB	6/14/2019 13:05	SMPL
K3208-04	MH0BD8	24989	4.7762 -		ISM02.4	PPB	6/14/2019 13:07	SMPL
K3208-05	MH0BD9	32067	6.1306 -		ISM02.4	PPB	6/14/2019 13:10	SMPL
K3208-06	MH0BE0	7198	1.3718 -		ISM02.4	PPB	6/14/2019 13:12	SMPL
K3208-07	MH0BE1	2313	0.4371 -		ISM02.4	PPB	6/14/2019 13:14	SMPL
K3208-08	MH0BE2	402	0.0714 -		ISM02.4	PPB	6/14/2019 13:17	SMPL
K3208-09	MH0BE3	745	0.137 -		ISM02.4	PPB	6/14/2019 13:19	SMPL
K3208-10	MH0BE4	2336	0.4415 -		ISM02.4	PPB	6/14/2019 13:21	SMPL
K3208-11	MH0BE5	5152	0.9803 -		ISM02.4	PPB	6/14/2019 13:24	SMPL
K3208-12	MH0BE6	3081	0.584 -		ISM02.4	PPB	6/14/2019 13:26	SMPL
K3208-13	MH0BE7	14907	2.847 -		ISM02.4	PPB	6/14/2019 13:28	SMPL
K3208-14	MH0BF5	7403	1.4111 -		ISM02.4	PPB	6/14/2019 13:30	SMPL
K3208-15	MH0BF6	13794	2.634 -		ISM02.4	PPB	6/14/2019 13:33	SMPL
K3208-16	MH0AW8	5545	1.0555 -		ISM02.4	PPB	6/14/2019 13:35	SMPL
K3208-17	MH0AX4	1664	0.3129 -		ISM02.4	PPB	6/14/2019 13:37	SMPL
K3208-18	MH0AX5	1469	0.2756 -		ISM02.4	PPB	6/14/2019 13:39	SMPL
K3208-19	MH0AX7	2173	0.4103 -		ISM02.4	PPB	6/14/2019 13:42	SMPL
CCV069	CCV069	24934	4.7657 -		ISM02.4	PPB	6/14/2019 13:47	SMPL
CCB069	CCB069	-12	-0.0078 -		ISM02.4	PPB	6/14/2019 13:50	SMPL
K3208-20	MH0AZ0	9714	1.8533 -		ISM02.4	PPB	6/14/2019 13:52	SMPL
K3208-21	MH0AZ0D	9480	1.8085 -		ISM02.4	PPB	6/14/2019 13:54	SMPL
K3208-22	MH0AZ0S	24899	4.759 -		ISM02.4	PPB	6/14/2019 14:00	SMPL
CCV070	CCV070	27747	5.304 -		ISM02.4	PPB	6/14/2019 14:02	SMPL
CCB070	CCB070	9	-0.0038 -		ISM02.4	PPB	6/14/2019 14:04	SMPL