



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103407

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/14/2019	09:50
2	K3369-02		1	8	100.00	>212.0	>212.0	06/14/2019	10:20
3	K3369-02DUP		1	8	100.00	>212.0	>212.0	06/14/2019	10:45
4	K3373-01		1	9	100.00	>212.0	>212.0	06/14/2019	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 103407

WorkList Name : FP61419

WorkList ID : 127431

Date : 06-14-2019 09:24:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2019	Water	K3369-02	C Flash Point	Cool 4 deg C	PSEG01	O31	FRAC-TANK-COMP	06/13/2019	1010A
06/20/2019	Water	K3373-01	F Flash Point	Cool 4 deg C	PSEG01	O31	13857-13964-13981	06/13/2019	1010A

Date/Time 06/14/19 09:25
 Received by: KS
 Relinquished by: CP

Date/Time 06/14/19 11:15
 Received by: CP
 Relinquished by: KS