



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 6/18/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:25
In Date: 06/14/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 06/15/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103426

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3154-01	JC-02-061419-A	1	1.11	9.8	8.97	90.4
K3154-03	JC-02-061419-B	2	1.11	9.82	9.04	91.0
K3154-05	JC-02-061419-C	3	1.1	9.63	8.75	89.7
K3154-07	JC-02-061419-D	4	1.09	9.65	8.92	91.5
K3154-09	JC-02-061419-E	5	1.09	9.58	8.66	89.2
K3154-11	JC-02-061419-F	6	1.11	9.58	8.65	89.0
K3154-13	JC-02-061419-G	7	1.09	9.68	9.02	92.3
K3366-05	BRANCH-LEAVES	8	1.00	2.00	2.00	100.0
K3376-01	NEW-DOVER-TP	9	1.08	9.81	8.67	86.9
K3376-02	NEW-DOVER-TP-1FT	10	1.08	9.81	8.67	86.9
K3376-03	NEW-DOVER-TP	11	1.08	9.81	8.67	86.9
K3382-01	CONCRETE	12	1.00	2.00	2.00	100.0
K3382-05	GRAVEL-1	13	1.00	2.00	2.00	100.0
K3382-07	GRAVEL-2	14	1.00	2.00	2.00	100.0
K3384-05	SVOC-GPC-BLANK	15	1.00	2.00	2.00	100.0
K3384-06	PEST-GPC-BLANK	16	1.00	2.00	2.00	100.0
K3384-07	PEST-GPC-BLANK-SPIKE	17	1.00	2.00	2.00	100.0
K3384-08	PCB-GPC-BLANK	18	1.00	2.00	2.00	100.0
K3384-09	PCB-GPC-BLANK-SPIKE	19	1.00	2.00	2.00	100.0
K3384-10	SVOC-GPC2-BLANK	20	1.00	2.00	2.00	100.0
K3384-11	PEST-GPC2-BLANK	21	1.00	2.00	2.00	100.0
K3384-12	PEST-GPC2-BLANK-SPIKE	22	1.00	2.00	2.00	100.0
K3384-13	PCB-GPC2-BLANK	23	1.00	2.00	2.00	100.0
K3384-14	PCB-GPC2-BLANK -SPIKE	24	1.00	2.00	2.00	100.0
K3389-01	GS-SOIL-PILE	25	1.09	9.64	8.58	87.6
K3389-02	GS-SOIL-PILE-A	26	1.09	9.64	8.58	87.6
K3389-03	GS-SOIL-PILE-B	27	1.09	9.64	8.58	87.6
K3389-04	GS-SOIL-PILE-C	28	1.09	9.64	8.58	87.6
K3389-05	GS-SOIL-PILE-D	29	1.09	9.64	8.58	87.6
K3389-06	GS-SOIL-PILE-E	30	1.09	9.64	8.58	87.6
K3389-07	GS-SOIL-PILE-G	31	1.09	9.64	8.58	87.6
K3389-08	GS-SOIL-PILE	32	1.09	9.64	8.58	87.6
K3393-01	DRILLING-MUD	33	1.09	9.81	6.33	60.1
K3396-01	390612709.2	34	1.18	9.98	1.2	0.2
K3397-01	390610038.0	35	1.14	9.81	5.91	55.0
K3398-01	390611024.0	36	1.12	9.89	1.7	6.6
K3400-01	390610037.0	37	1.14	9.97	1.73	6.7
K3406-01	BAT-B14	38	1.13	9.89	8.8	87.6
K3406-02	BAT-B14	39	1.18	9.52	8.42	86.8



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 6/18/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:25
In Date: 06/14/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 06/15/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103426

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3406-03	BAT-B14-GEO	40	1.15	9.95	8.65	85.2
K3406-04	BAT-B14-25-27	41	1.17	9.92	7.46	71.9
K3407-02	OILY-DEBRIS	42	1.00	2.00	2.00	100.0
K3407-05	OIL	43	1.00	2.00	2.00	100.0
K3409-01	T-1	44	1.1	9.89	8.53	84.5
K3409-02	T-1A	45	1.19	9.98	8.75	86.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$