

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB103429

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

| Reagent/Standard | Lot/Log # |
|-----------------------------------|-----------|
| Sodium thiosulfate titrant 0.025N | WP73671 |
| IODINE SOLUTION .025N 1L | W2480 |
| Starch Solution, 4L | W2379 |

| Seq | Lab ID | True Value (mg/l) | DF | Initial Volume (mL) | Final Volume (ml) | T1 (ml) | T2 Initial | T2 Final | T2 Diff. (ml) | T1 - T2 Diff (mL) | Value Corrected With Blank | Result (ppm) | Anal Date | Anal Time |
|-----|-------------|-------------------|----|---------------------|-------------------|---------|------------|----------|---------------|-------------------|----------------------------|--------------|------------|-----------|
| 1 | PB120637BL | | 1 | 50 | 50 | 2.00 | 0.00 | 1.84 | 1.84 | 0.16 | 0.00 | 0.00 | 06/14/2019 | 14:45 |
| 2 | K3369-02 | | 1 | 50 | 50 | 2.00 | 1.84 | 3.64 | 1.80 | 0.20 | 0.04 | 0.32 | 06/14/2019 | 14:48 |
| 3 | K3369-02DUP | | 1 | 50 | 50 | 2.00 | 3.64 | 5.46 | 1.82 | 0.18 | 0.02 | 0.16 | 06/14/2019 | 14:52 |

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume