



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 6/18/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 16:00
In Date: 06/17/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 07:30
Out Date: 06/18/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103450

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3390-01	MH0CC4	1	1.14	9.9	9.57	96.2
K3390-02	MH0CC5	2	1.15	9.64	8.98	92.2
K3390-03	MH0CC6	3	1.14	9.68	9.18	94.1
K3390-04	MH0CC7	4	1.15	9.58	9.38	97.6
K3390-05	MH0CC7D	5	1.15	9.58	9.38	97.6
K3390-06	MH0CC7S	6	1.15	9.58	9.38	97.6
K3390-07	MH0CC8	7	1.13	9.69	9.52	98.0
K3390-08	MH0CC9	8	1.13	9.77	9.61	98.1
K3390-09	MH0CD0	9	1.14	9.89	9.57	96.3
K3390-10	MH0CD1	10	1.13	9.81	9.54	96.9
K3390-11	MH0CD2	11	1.12	9.86	9.44	95.2
K3390-12	MH0CD3	12	1.12	9.91	9.63	96.8
K3390-13	MH0CD4	13	1.13	9.73	8.9	90.3
K3390-14	MH0CD5	14	1.13	9.83	9.67	98.2
K3390-15	MH0CD6	15	1.13	9.72	9.36	95.8
K3390-16	MH0CD7	16	1.13	9.75	9.36	95.5
K3390-17	MH0CD8	17	1.13	9.83	9.34	94.4
K3390-18	MH0CE0	18	1.12	9.62	9.4	97.4
K3390-19	MH0CE1	19	1.14	9.7	9.47	97.3
K3390-20	MH0CE2	20	1.13	9.74	9.61	98.5
K3390-21	MH0CE3	21	1.12	9.6	9.2	95.3
K3390-22	MH0CE4	22	1.12	9.56	9.13	94.9

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$