



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 6/18/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 06/17/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 08:11
Out Date: 06/18/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103456

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3157-01	CL-01-061719-A	1	1.14	9.72	9.08	92.5
K3157-03	CL-01-061719-B	2	1.13	9.63	9.14	94.2
K3157-05	CL-01-061719-C	3	1.15	9.64	9.2	94.8
K3157-07	CL-01-061719-D	4	1.19	9.79	9.27	94.0
K3157-09	CL-01-061719-E	5	1.14	9.74	8.98	91.2
K3158-03	OK-01-061719	6	1.13	9.66	9.19	94.5
K3163-09	SU-03-061719-A	7	1.11	9.79	9.13	92.4
K3163-11	SU-03-061719-B	8	1.15	9.95	9.22	91.7
K3382-06	GRAVEL-1	9	1.00	2.00	2.00	100.0
K3382-08	GRAVEL-2	10	1.00	2.00	2.00	100.0
K3417-01	CONTROL-HOUSE-BRICKS	11	1.00	2.00	2.00	100.0
K3418-01	71630	12	1.00	2.00	2.00	100.0
K3419-01	TP-A	13	1.13	9.76	9.31	94.8
K3419-02	TP-A	14	1.13	9.76	9.31	94.8
K3419-03	TP-B	15	1.15	9.52	7.37	74.3
K3419-04	TP-B	16	1.15	9.52	7.37	74.3
K3419-05	TP-C	17	1.19	9.85	8.32	82.3
K3419-06	TP-C	18	1.19	9.85	8.32	82.3
K3419-07	TP-D	19	1.16	9.97	9.61	95.9
K3419-08	TP-D	20	1.16	9.97	9.61	95.9
K3419-09	TP-E	21	1.17	9.98	9.65	96.3
K3419-10	TP-E	22	1.17	9.98	9.65	96.3
K3420-01	BAT-B12	23	1.16	9.57	8.72	89.9
K3420-02	BAT-B12	24	1.19	9.89	9.12	91.1
K3421-01	SAMPLE-1	25	1.15	9.85	8.89	89.0
K3421-02	SAMPLE-1	26	1.15	9.85	8.89	89.0
K3421-03	SAMPLE-1A	27	1.15	9.85	8.89	89.0
K3421-05	SAMPLE-2	28	1.14	9.98	8.92	88.0
K3421-06	SAMPLE-2	29	1.14	9.98	8.92	88.0
K3421-07	SAMPLE-2A	30	1.14	9.98	8.92	88.0
K3422-01	390531083.3	31	1.12	9.97	1.36	2.7

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$