



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103457

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/17/2019	18:00
2	CAL2	1	Water	NA	NA	20.1	6.99	06/17/2019	18:01
3	CAL3	1	Water	NA	NA	20.1	10.03	06/17/2019	18:02
4	ICV	1	Water	NA	NA	20.2	7.00	06/17/2019	18:04
5	CCV1	1	Water	NA	NA	20.1	2.02	06/17/2019	18:05
6	K3157-02	1	Solid	20.03	20	20.6	8.45	06/17/2019	18:06
7	K3157-04	1	Solid	20.02	20	21.0	8.96	06/17/2019	18:07
8	K3157-06	1	Solid	20.03	20	22.2	8.75	06/17/2019	18:10
9	K3157-08	1	Solid	20.01	20	22.2	9.28	06/17/2019	18:11
10	K3157-10	1	Solid	20.03	20	22.4	8.55	06/17/2019	18:12
11	K3158-04	1	Solid	20.05	20	22.2	9.93	06/17/2019	18:15
12	K3163-10	1	Solid	20.03	20	22.2	8.52	06/17/2019	18:16
13	K3163-12	1	Solid	20.04	20	22.0	8.35	06/17/2019	18:20
14	K3421-04	1	Solid	20.04	20	21.8	5.58	06/17/2019	18:21
15	K3421-08	1	Solid	20.02	20	20.7	5.56	06/17/2019	18:22
16	CCV2	1	Water	NA	NA	20.1	12.02	06/17/2019	18:24
17	K3421-08DUP	1	Solid	20.01	20	21.6	5.55	06/17/2019	18:25
18	CCV3	1	Water	NA	NA	20.2	2.01	06/17/2019	18:26

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k3158

WorkList ID : 127519

Date : 06-17-2019 13:22:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/17/2019	Solid	K3157-02	Corrosivity	Cool 4 deg C	PSEG05	O31	CL-01-061719-A	06/17/2019	9045D
06/17/2019	Solid	K3157-04	Corrosivity	Cool 4 deg C	PSEG05	O31	CL-01-061719-B	06/17/2019	9045D
06/17/2019	Solid	K3157-06	Corrosivity	Cool 4 deg C	PSEG05	O31	CL-01-061719-C	06/17/2019	9045D
06/17/2019	Solid	K3157-08	Corrosivity	Cool 4 deg C	PSEG05	O31	CL-01-061719-D	06/17/2019	9045D
06/17/2019	Solid	K3157-10	Corrosivity	Cool 4 deg C	PSEG05	O31	CL-01-061719-E	06/17/2019	9045D
06/17/2019	Solid	K3158-04	Corrosivity	Cool 4 deg C	PSEG05		OK-01-061719	06/17/2019	9045D
06/17/2019	Solid	K3163-10	Corrosivity	Cool 4 deg C	PSEG05		SU-03-061719-A	06/17/2019	9045D
06/17/2019	Solid	K3163-12	Corrosivity	Cool 4 deg C	PSEG05		SU-03-061719-B	06/17/2019	9045D
06/17/2019	Solid	K3421-04	Corrosivity	Cool 4 deg C	PSEG01	K11	SAMPLE-1	06/17/2019	9045D
06/17/2019	Solid	K3421-08	Corrosivity	Cool 4 deg C	PSEG01	K11	SAMPLE-2	06/17/2019	9045D

Date/Time 06-17-19 16:00
 Received by: Jp
 Relinquished by: Cb

Date/Time 06-17-19 18:10
 Received by: Cb
 Relinquished by: Jp