



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103458

Ambient Barometric Pressure (mmHg): 755.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.00	80.6	80.9	06/17/2019	11:58
2	K3382-03		1	7	100.00	>212.0	>212.0	06/17/2019	12:20
3	K3382-03DUP		1	7	100.00	>212.0	>212.0	06/17/2019	12:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP61719

WorkList ID : 127496

Date : 06-17-2019 09:20:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/28/2019	Water	K3382-03	F	Cool 4 deg C	PSEG01	O31	SUMP	06/14/2019	1010A

Date/Time 06/17/19 11:30
 Received by: KS
 Relinquished by: CP

Date/Time 06/17/19 12:55
 Received by: CP
 Relinquished by: KS