

OVEN TEMP1: 103 °C TEMP: 105 °C

DATE & TIME IN: 06/14/2019 15:20 OUT: 06/17/2019 08:30

OVEN TEMP2: 105 °C TEMP: 105 °C

DATE & TIME IN: 06/17/2019 10:00 OUT: 06/17/2019 11:40

Run Number: LB103464

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Dish + Sample Weight (g)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result %
1	LB103464BL	LB103464BL	89.4579	89.4579	100	89.4579	89.4579	89.4579	0	0
2	K3352-01	20190612-LOC-11	68.8433	68.8433	100	68.8605	68.8605	68.8605	0.0172	172
3	K3352-01DUP	20190612-LOC-11DUP	85.1002	85.1002	100	85.1175	85.1175	85.1175	0.0173	173
4	K3352-03	20190612-LOC-12	90.011	90.011	100	90.0161	90.0161	90.0161	0.0051	51

A = Final Empty Dish Weight (g)

B = Dish + Sample Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result \%} = (C - A) * 100 / (B - A)$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS061419

WorkList ID : 127445

Date : 06-14-2019 14:24:50

AB1034624

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/18/2019	Solid	K3352-01	ASTM TS	Cool 4 deg C	SEVE02	H61	20190612-LOC-11	06/12/2019	SM2540 B
06/18/2019	Solid	K3352-03	ASTM TS	Cool 4 deg C	SEVE02	H61	20190612-LOC-12	06/12/2019	SM2540 B

Date/Time 06/14/19, 14:40
 Received by: AS
 Relinquished by: CS

Date/Time 06/14/19, 15:50
 Received by: AS
 Relinquished by: AS