



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103477

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/14/2019	15:15
2	CAL2	1	Water	NA	NA	20.2	7.00	06/14/2019	15:16
3	CAL3	1	Water	NA	NA	20.2	10.03	06/14/2019	15:17
4	ICV	1	Water	NA	NA	20.1	7.01	06/14/2019	15:19
5	CCV1	1	Water	NA	NA	20.1	2.02	06/14/2019	15:20
6	K3406-02	1	Solid	20.03	20	21.9	10.82	06/14/2019	15:22
7	K3406-03	1	Solid	20.02	20	21.8	8.29	06/14/2019	15:24
8	K3406-03DUP	1	Solid	20.03	20	21.7	8.30	06/14/2019	15:25
9	CCV2	1	Water	NA	NA	20.1	12.02	06/14/2019	15:26

AB 103477

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3406

WorkList ID : 127571

Date : 06-14-2019 15:02:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/14/2019	Solid	K3406-02	~X pH	Cool 4 deg C	loui01	H11	BAT-B14	06/14/2019	9045D
06/14/2019	Solid	K3406-03	~X pH	Cool 4 deg C	loui01	H11	BAT-B14-GEO	06/14/2019	9045D

Date/Time 06-18-19 15:10
Received by: JF
Relinquished by: R.

Date/Time 06-18-19 16:23
Received by: PC
Relinquished by: Jb