



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 6/19/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 16:00
In Date: 06/18/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 06/19/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103478

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3404-01	MH0CB8	1	1.14	9.8	8.64	86.6
K3404-02	MH0CB9	2	1.16	9.55	8.39	86.2
K3404-03	MH0CC0	3	1.18	9.9	9.32	93.3
K3404-04	MH0CC0D	4	1.18	9.9	9.32	93.3
K3404-05	MH0CC0S	5	1.18	9.9	9.32	93.3
K3404-06	MH0CC1	6	1.18	9.74	9.29	94.7
K3404-07	MH0CC2	7	1.1	9.74	9.15	93.2
K3404-08	MH0BU7	8	1.17	9.86	9.61	97.1
K3404-09	MH0BZ1	9	1.19	9.72	8.82	89.4
K3404-10	MH0CB2	10	1.12	9.71	9.26	94.8
K3404-11	MH0CC3	11	1.18	9.95	9.48	94.6
K3404-12	MH0BT9	12	1.15	9.6	8.88	91.5
K3404-13	MH0BU0	13	1.15	9.95	9.43	94.1
K3404-14	MH0BU1	14	1.17	9.96	9.32	92.7
K3404-15	MH0BU2	15	1.17	9.78	9.3	94.4
K3404-16	MH0BU3	16	1.17	9.7	9.14	93.4
K3404-17	MH0BU4	17	1.18	9.96	8.51	83.5
K3404-18	MH0BU5	18	1.14	9.58	8.74	90.0
K3404-20	MH0BV3	19	1.1	9.6	9.3	96.5
K3404-21	MH0BV4	20	1.18	9.8	9.22	93.3
K3404-22	MH0BV5	21	1.2	9.98	8.77	86.2

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$