



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103490

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/18/2019	16:00
2	CAL2	1	Water	NA	NA	20.1	7.00	06/18/2019	16:01
3	CAL3	1	Water	NA	NA	20.1	10.02	06/18/2019	16:02
4	ICV	1	Water	NA	NA	20.1	7.01	06/18/2019	16:04
5	CCV1	1	Water	NA	NA	20.2	2.02	06/18/2019	16:05
6	K3429-02	1	Water	NA	NA	20.7	8.36	06/18/2019	16:06
7	K3433-01	1	Water	NA	NA	22.5	6.31	06/18/2019	16:08
8	K3434-01	1	Water	NA	NA	22.7	8.92	06/18/2019	16:10
9	K3441-05	1	Water	NA	NA	21.6	6.77	06/18/2019	16:11
10	K3441-05DUP	1	Water	NA	NA	21.7	6.78	06/18/2019	16:12
11	CCV2	1	Water	NA	NA	20.2	12.01	06/18/2019	16:15

103496

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3441 W

WorkList ID : 127577

Date : 06-18-2019 15:31:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/17/2019	Water	K3429-02	pH	Cool 4 deg C	TETR06	J11	TT-002-IDWGW-20190617	06/17/2019	9040C
06/18/2019	Water	K3433-01	pH	Cool 4 deg C	PSEG01	K12	OILY-WATER	06/18/2019	9040C
06/18/2019	Water	K3434-01	pH	Cool 4 deg C	PSEG01	K12	OILY-WATER	06/18/2019	9040C
06/18/2019	Water	K3441-05	pH	Cool 4 deg C	PSEG01	K13	GWMH-4	06/18/2019	9040C

Date/Time 06-18-19 15:35
 Received by: JF PC
 Relinquished by: JF

Date/Time 06-18-19 17:35
 Received by: PC
 Relinquished by: JF