



Analytical Summary Report

Analysis Method: 9041A

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103491

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467
0.0-3.0 pH paper refill	W1931
Test Paper, PH, Short Range 6.0/8.0	W2350
Test Paper,PH Short Range 9.0/10.0	W1939
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.0	06/18/2019	16:00
2	CAL2	1	Water	NA	NA	20.1	7.0	06/18/2019	16:01
3	CAL3	1	Water	NA	NA	20.1	10.0	06/18/2019	16:02
4	ICV	1	Water	NA	NA	20.1	7.0	06/18/2019	16:04
5	CCV1	1	Water	NA	NA	20.2	2.0	06/18/2019	16:30
6	K3434-02	1	Solid	20.02	20	22.2	5.2	06/18/2019	16:31
7	K3434-02DUP	1	Solid	20.01	20	22.1	5.2	06/18/2019	16:33
8	CCV2	1	Water	NA	NA	20.1	12.0	06/18/2019	16:35

WORKLIST(Hardcopy Internal Chain)

2B 103491

WorkList Name : PH K3434 S

WorkList ID : 127581

Date : 06-18-2019 15:37:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/18/2019	Solid	K3434-02	pH	Cool 4 deg C	PSEG01	K12	OIL	06/18/2019	9041A

Date/Time 06-18-19 15:40
 Received by: js
 Relinquished by: pc

Date/Time 06-18-19 17:35
 Received by: pc
 Relinquished by: js