



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103492

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/18/2019	16:00
2	CAL2	1	Water	NA	NA	20.1	7.00	06/18/2019	16:01
3	CAL3	1	Water	NA	NA	20.1	10.02	06/18/2019	16:02
4	ICV	1	Water	NA	NA	20.1	7.01	06/18/2019	16:04
5	CCV1	1	Water	NA	NA	20.2	2.02	06/18/2019	16:40
6	K3441-04	1	Solid	20.03	20	21.7	7.16	06/18/2019	16:41
7	K3441-04DUP	1	Solid	20.04	20	21.8	7.17	06/18/2019	16:42
8	CCV2	1	Water	NA	NA	20.1	12.02	06/18/2019	16:44

WorkList Name : corrosivity k3432

WorkList ID : 127575

Date : 06-18-2019 15:30:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/18/2019	Solid	K3441-04	Corrosivity	Cool 4 deg C	PSEG01	K13	SAMPLE-3	06/18/2019	9045D

Date/Time 06-18-19 15:30
Received by: JH
Relinquished by: PC

Date/Time 06-18-19 17:30
Received by: PC
Relinquished by: JP