


7470A

LB103497

INSTRUMENT ID : CU1



Linear

A= 0.0000e+000

B= 2.7622e-004

C= -1.3397e-001

Rho= 0.9973549

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	462	0.000	462				
0.2	0.200	0.171	-0.029	1104	0.0 %	1104				
2.5	2.500	2.447	-0.053	9343	0.0 %	9343				
5.0	5.000	4.979	-0.021	18510	0.0 %	18510				
7.5	7.500	8.014	0.514	29498	0.0 %	29498				
10.0	10.000	9.596	-0.404	35225	0.0 %	35225				

LB103497 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 19 Jun 2019 10:46:20					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	500	462	-	0.0000	7470A	PPB	19 Jun 2019 11:03:08	S	Std
0.2 - 1	502	1104	-	0.2000	7470A	PPB	19 Jun 2019 11:05:16	S	Std
2.5 - 1	525	9343	-	2.5000	7470A	PPB	19 Jun 2019 11:07:25	S	Std
5.0 - 1	550	18510	-	5.0000	7470A	PPB	19 Jun 2019 11:09:33	S	Std
7.5 - 1	575	29498	-	7.5000	7470A	PPB	19 Jun 2019 11:11:41	S	Std
10.0 - 1	5100	35225	-	10.0000	7470A	PPB	19 Jun 2019 11:18:11	S	Std
ICV03 - 1	ICV03	15538	4.1579	-	7470A	PPB	19 Jun 2019 11:24:40	U	SMPL
ICB03 - 1	ICB03	195	-0.0801	-	7470A	PPB	19 Jun 2019 11:26:46	U	SMPL
CCV95 - 1	CCV95	18006	4.8396	-	7470A	PPB	19 Jun 2019 11:28:53	U	SMPL
CCB95 - 1	CCB95	399	-0.0238	-	7470A	PPB	19 Jun 2019 11:31:00	U	SMPL
CRA - 1	CRA	1090	0.1671	-	7470A	PPB	19 Jun 2019 11:33:06	U	SMPL
HighStd - 1	HighStd	35708	9.7293	-	7470A	PPB	19 Jun 2019 11:35:12	U	SMPL
ChkStd - 1	ChkStd	25612	6.9406	-	7470A	PPB	19 Jun 2019 11:37:19	U	SMPL
PB120757BL - 1	PBW	182	-0.0837	-	7470A	PPB	19 Jun 2019 11:39:27	U	SMPL
PB120757BS - 1	LCSW	14422	3.8497	-	7470A	PPB	19 Jun 2019 11:41:35	U	SMPL
K3157-02 - 1	CL-01-061719-A	184	-0.0831	-	7470A	PPB	19 Jun 2019 11:43:42	U	SMPL
K3157-02DUP - 1	CL-01-061719-ADUP	207	-0.0768	-	7470A	PPB	19 Jun 2019 11:45:50	U	SMPL
K3157-02MS - 1	CL-01-061719-AMS	15211	4.0676	-	7470A	PPB	19 Jun 2019 11:47:58	U	SMPL
K3157-02MSD - 1	CL-01-061719-AMSD	15617	4.1797	-	7470A	PPB	19 Jun 2019 11:50:06	U	SMPL
K3157-04 - 1	CL-01-061719-B	207	-0.0768	-	7470A	PPB	19 Jun 2019 11:52:13	U	SMPL
CCV96 - 1	CCV96	18132	4.8744	-	7470A	PPB	19 Jun 2019 11:54:19	U	SMPL
CCB96 - 1	CCB96	206	-0.0771	-	7470A	PPB	19 Jun 2019 11:56:25	U	SMPL
K3157-06 - 1	CL-01-061719-C	295	-0.0525	-	7470A	PPB	19 Jun 2019 11:58:31	U	SMPL
K3157-08 - 1	CL-01-061719-D	247	-0.0657	-	7470A	PPB	19 Jun 2019 12:00:37	U	SMPL
K3157-10 - 1	CL-01-061719-E	244	-0.0666	-	7470A	PPB	19 Jun 2019 12:02:45	U	SMPL
K3158-04 - 1	OK-01-061719	301	-0.0508	-	7470A	PPB	19 Jun 2019 12:04:51	U	SMPL
K3163-10 - 1	SU-03-061719-A	296	-0.0522	-	7470A	PPB	19 Jun 2019 12:06:58	U	SMPL
K3163-12 - 1	SU-03-061719-B	356	-0.0356	-	7470A	PPB	19 Jun 2019 12:09:06	U	SMPL
K3417-01 - 1	CONTROL-HOUSE-BRICKS	345	-0.0387	-	7470A	PPB	19 Jun 2019 12:11:13	U	SMPL
K3421-04 - 1	SAMPLE-1	318	-0.0461	-	7470A	PPB	19 Jun 2019 12:13:21	U	SMPL
K3421-08 - 1	SAMPLE-2	334	-0.0417	-	7470A	PPB	19 Jun 2019 12:15:29	U	SMPL
PB120758BL - 1	PBW	249	-0.0652	-	7470A	PPB	19 Jun 2019 12:17:36	U	SMPL
CCV97 - 1	CCV97	17838	4.7932	-	7470A	PPB	19 Jun 2019 12:19:42	U	SMPL
CCB97 - 1	CCB97	491	0.0017	-	7470A	PPB	19 Jun 2019 12:21:49	U	SMPL
PB120758BS - 1	LCSW	14440	3.8546	-	7470A	PPB	19 Jun 2019 12:23:56	U	SMPL
K3382-02 - 1	SUMP	1180	0.1920	-	7470A	PPB	19 Jun 2019 12:26:03	U	SMPL
K3382-02DUP - 1	SUMPDUP	1340	0.2362	-	7470A	PPB	19 Jun 2019 12:28:09	U	SMPL
K3382-02MS - 1	SUMPMS	15937	4.2681	-	7470A	PPB	19 Jun 2019 12:30:16	U	SMPL
K3382-02MSD - 1	SUMPMDS	14655	3.9140	-	7470A	PPB	19 Jun 2019 12:32:24	U	SMPL
K3429-02 - 1	TT-002-IDWGW-20190617	1399	0.2525	-	7470A	PPB	19 Jun 2019 12:34:32	U	SMPL
K3433-01 - 1	OILY-WATER	13	-0.1304	-	7470A	PPB	19 Jun 2019 12:36:40	U	SMPL
K3441-05 - 1	GWMH-4	4175	1.0192	-	7470A	PPB	19 Jun 2019 12:38:48	U	SMPL
PB120757TB - 1	PB120698TB	286	-0.0550	-	7470A	PPB	19 Jun 2019 12:40:57	U	SMPL
K3157-02LX5 - 1		-76	-0.1550	-	7470A	PPB	19 Jun 2019 12:43:04	U	SMPL
CCV98 - 1	CCV98	17824	4.7894	-	7470A	PPB	19 Jun 2019 12:45:11	U	SMPL
CCB98 - 1	CCB98	221	-0.0729	-	7470A	PPB	19 Jun 2019 12:47:18	U	SMPL
K3157-02A - 1		15311	4.0952	-	7470A	PPB	19 Jun 2019 12:49:25	U	SMPL
K3382-02LX5 - 1		66	-0.1157	-	7470A	PPB	19 Jun 2019 12:51:32	U	SMPL
K3382-02A - 1		15466	4.1380	-	7470A	PPB	19 Jun 2019 12:53:38	U	SMPL
CCV99 - 1	CCV99	17860	4.7993	-	7470A	PPB	19 Jun 2019 12:55:45	U	SMPL
CCB99 - 1	CCB99	354	-0.0362	-	7470A	PPB	19 Jun 2019 12:57:53	U	SMPL