



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103507

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/19/2019	10:40
2	K3433-02		1	9	100.00	>212.0	>212.0	06/19/2019	11:05
3	K3433-02DUP		1	8	100.00	>212.0	>212.0	06/19/2019	11:35
4	K3441-06		1	8	100.00	>212.0	>212.0	06/19/2019	11:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP61919

WorkList ID : 127630

Date : 06-19-2019 09:23:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/25/2019	Water	K3433-02	Flash Point	Cool 4 deg C	PSEG01	K12	OILY-WATER	06/18/2019	1010A
07/01/2019	Water	K3441-06	Flash Point	Cool 4 deg C	PSEG01	K13	GWMH-4	06/18/2019	1010A

Date/Time 06/19/19 10:15
 Received by: KS
 Relinquished by: KS

Date/Time 06/19/19 12:10
 Received by: KS
 Relinquished by: KS