


LB103509

ISM02.4

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 2.3001e-004

C= 1.7147e-003

Rho= 0.9998419

Accept=Accepted

slope
y intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.041	0.041	171	0.000	171				
0.05	0.050									
0.20	0.200	0.212	0.012	913	0.0 %	913				
2.50	2.500	2.528	0.028	10983	0.0 %	10983				
5.00	5.000	4.862	-0.138	21133	0.0 %	21133				
7.50	7.500	7.492	-0.008	32565	0.0 %	32565				
10.0	10.000	10.065	0.065	43753	0.0 %	43753				

%D
—
6
1
3
0
1

RS 06-19-19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	171	-		0 ISM02.4	PPB	6/19/2019 14:23	Std
	0.2 S0.2	913	-		0.2 ISM02.4	PPB	6/19/2019 14:26	Std
	2.5 S2.5	10983	-		2.5 ISM02.4	PPB	6/19/2019 14:28	Std
	5 S5.0	21133	-		5 ISM02.4	PPB	6/19/2019 14:30	Std
	7.5 S7.5	32565	-		7.5 ISM02.4	PPB	6/19/2019 14:33	Std
	10 S10.0	43753	-		10 ISM02.4	PPB	6/19/2019 14:35	Std
ICV	ICV	18613	4.2828 -		ISM02.4	PPB	6/19/2019 14:40	SMPL
ICB	ICB	75	0.019 -		ISM02.4	PPB	6/19/2019 14:42	SMPL
CCV01	CCV101	21124	4.8603 -		ISM02.4	PPB	6/19/2019 14:44	SMPL
CCB01	CCB01	32	0.0091 -		ISM02.4	PPB	6/19/2019 14:46	SMPL
PB120770BL	PBS770	217	0.0516 -		ISM02.4	PPB	6/19/2019 14:49	SMPL
K3390-01	MH0CC4	3605	0.8309 -		ISM02.4	PPB	6/19/2019 14:51	SMPL
K3390-02	MH0CC5	327	0.0769 -		ISM02.4	PPB	6/19/2019 14:53	SMPL
K3390-03	MH0CC6	3595	0.8286 -		ISM02.4	PPB	6/19/2019 14:55	SMPL
K3390-04	MH0CC7	9302	2.1412 -		ISM02.4	PPB	6/19/2019 14:58	SMPL
K3390-05	MH0CC7D	9329	2.1474 -		ISM02.4	PPB	6/19/2019 15:00	SMPL
K3390-06	MH0CC7S	21329	4.9075 -		ISM02.4	PPB	6/19/2019 15:07	SMPL
K3390-07	MH0CC8	9968	2.2944 -		ISM02.4	PPB	6/19/2019 15:09	SMPL
K3390-08	MH0CC9	8715	2.0062 -		ISM02.4	PPB	6/19/2019 15:11	SMPL
K3390-09	MH0CD0	3287	0.7577 -		ISM02.4	PPB	6/19/2019 15:14	SMPL
K3390-10	MH0CD1	3623	0.835 -		ISM02.4	PPB	6/19/2019 15:16	SMPL
K3390-11	MH0CD2	1017	0.2356 -		ISM02.4	PPB	6/19/2019 15:18	SMPL
K3390-12	MH0CD3	4539	1.0457 -		ISM02.4	PPB	6/19/2019 15:20	SMPL
K3390-13	MH0CD4	23813	5.4788 -		ISM02.4	PPB	6/19/2019 15:23	SMPL
K3390-14	MH0CD5	9927	2.285 -		ISM02.4	PPB	6/19/2019 15:25	SMPL
K3390-15	MH0CD6	7882	1.8146 -		ISM02.4	PPB	6/19/2019 15:27	SMPL
K3390-16	MH0CD7	6545	1.5071 -		ISM02.4	PPB	6/19/2019 15:29	SMPL
K3390-17	MH0CD8	8963	2.0633 -		ISM02.4	PPB	6/19/2019 15:32	SMPL
K3390-18	MH0CE0	9747	2.2436 -		ISM02.4	PPB	6/19/2019 15:34	SMPL
K3390-19	MH0CE1	11216	2.5815 -		ISM02.4	PPB	6/19/2019 15:36	SMPL
CCV02	CCV02	20875	4.8031 -		ISM02.4	PPB	6/19/2019 15:39	SMPL
CCB02	CCB02	75	0.019 -		ISM02.4	PPB	6/19/2019 15:41	SMPL
K3390-20	MH0CE2	16428	3.7802 -		ISM02.4	PPB	6/19/2019 15:43	SMPL
K3390-21	MH0CE3	830	0.1926 -		ISM02.4	PPB	6/19/2019 15:45	SMPL
K3390-22	MH0CE4	2788	0.643 -		ISM02.4	PPB	6/19/2019 15:48	SMPL
PB120771BL	PBS771	10	0.004 -		ISM02.4	PPB	6/19/2019 15:50	SMPL
K3391-01	MH0CD9	4756	1.0956 -		ISM02.4	PPB	6/19/2019 15:52	SMPL
K3391-02	MH0CD9D	4755	1.0954 -		ISM02.4	PPB	6/19/2019 15:54	SMPL
K3391-03	MH0CD9S	16947	3.8996 -		ISM02.4	PPB	6/19/2019 15:57	SMPL
K3391-04	MH0CE5	6647	1.5306 -		ISM02.4	PPB	6/19/2019 15:59	SMPL
K3391-05	MH0CE6	27896	6.4179 -		ISM02.4	PPB	6/19/2019 16:01	SMPL
K3391-06	MH0CE7	26487	6.0939 -		ISM02.4	PPB	6/19/2019 16:04	SMPL
K3391-07	MH0CE8	21307	4.9024 -		ISM02.4	PPB	6/19/2019 16:06	SMPL
K3391-08	MH0CE9	25161	5.7889 -		ISM02.4	PPB	6/19/2019 16:08	SMPL
K3391-09	MH0CF1	1508	0.3486 -		ISM02.4	PPB	6/19/2019 16:10	SMPL
K3391-10	MH0CF2	22607	5.2014 -		ISM02.4	PPB	6/19/2019 16:13	SMPL

INSTRUMENT ID : CV1

K3391-11	MHOCF3	2785	0.6423 -	ISM02.4	PPB	6/19/2019 16:15	SMPL
K3391-12	MHOCF4	3408	0.7856 -	ISM02.4	PPB	6/19/2019 16:17	SMPL
K3391-13	MHOCF5	4836	1.114 -	ISM02.4	PPB	6/19/2019 16:20	SMPL
K3391-14	MHOCF6	5206	1.1991 -	ISM02.4	PPB	6/19/2019 16:22	SMPL
K3391-15	MHOCF7	43686	10.0497 -	ISM02.4	PPB	6/19/2019 16:24	SMPL
K3391-16	MHOCF8	18211	4.1903 -	ISM02.4	PPB	6/19/2019 16:26	SMPL
CCV03	CCV03	20478	4.7118 -	ISM02.4	PPB	6/19/2019 16:29	SMPL
CCB03	CCB03	39	0.0107 -	ISM02.4	PPB	6/19/2019 16:31	SMPL
K3391-17	MHOCF9	3234	0.7456 -	ISM02.4	PPB	6/19/2019 16:33	SMPL
K3391-18	MHOCG0	3838	0.8845 -	ISM02.4	PPB	6/19/2019 16:35	SMPL
K3391-19	MHOCG1	1909	0.4408 -	ISM02.4	PPB	6/19/2019 16:38	SMPL
K3391-20	MHOCG3	3986	0.9185 -	ISM02.4	PPB	6/19/2019 16:40	SMPL
K3391-21	MHOCG4	1345	0.3111 -	ISM02.4	PPB	6/19/2019 16:42	SMPL
K3391-22	MHOCG5	20046	4.6124 -	ISM02.4	PPB	6/19/2019 16:45	SMPL
PB120772BL	PBS772	-14	-0.0015 -	ISM02.4	PPB	6/19/2019 16:47	SMPL
K3399-01	MHOCF0	2088	0.482 -	ISM02.4	PPB	6/19/2019 16:49	SMPL
K3399-02	MHOCF0D	2102	0.4852 -	ISM02.4	PPB	6/19/2019 16:51	SMPL
K3399-03	MHOCF0S	13855	3.1884 -	ISM02.4	PPB	6/19/2019 16:58	SMPL
K3399-04	MH0AC4	8523	1.962 -	ISM02.4	PPB	6/19/2019 17:01	SMPL
K3399-05	MH0BU6	563	0.1312 -	ISM02.4	PPB	6/19/2019 17:03	SMPL
K3399-06	MH0BU8	2106	0.4861 -	ISM02.4	PPB	6/19/2019 17:05	SMPL
K3399-07	MH0BU9	3657	0.8428 -	ISM02.4	PPB	6/19/2019 17:08	SMPL
K3399-08	MH0BV0	4182	0.9636 -	ISM02.4	PPB	6/19/2019 17:10	SMPL
K3399-09	MH0BV1	4253	0.9799 -	ISM02.4	PPB	6/19/2019 17:12	SMPL
K3399-10	MH0BV2	6763	1.5572 -	ISM02.4	PPB	6/19/2019 17:14	SMPL
K3399-11	MH0BV8	1570	0.3628 -	ISM02.4	PPB	6/19/2019 17:17	SMPL
K3399-12	MH0BV9	4389	1.0112 -	ISM02.4	PPB	6/19/2019 17:19	SMPL
K3399-13	MH0BW0	4538	1.0455 -	ISM02.4	PPB	6/19/2019 17:21	SMPL
CCV04	CCV04	20196	4.6469 -	ISM02.4	PPB	6/19/2019 17:23	SMPL
CCB04	CCB04	39	0.0107 -	ISM02.4	PPB	6/19/2019 17:26	SMPL
K3399-14	MH0BW1	4571	1.0531 -	ISM02.4	PPB	6/19/2019 17:28	SMPL
K3399-15	MH0BW2	5080	1.1701 -	ISM02.4	PPB	6/19/2019 17:30	SMPL
K3399-16	MH0BW3	4605	1.0609 -	ISM02.4	PPB	6/19/2019 17:33	SMPL
K3399-17	MH0BW4	6597	1.5191 -	ISM02.4	PPB	6/19/2019 17:35	SMPL
K3399-18	MH0BW5	5609	1.2918 -	ISM02.4	PPB	6/19/2019 17:37	SMPL
K3399-19	MH0BW6	15533	3.5744 -	ISM02.4	PPB	6/19/2019 17:39	SMPL
K3399-20	MH0BW7	1657	0.3828 -	ISM02.4	PPB	6/19/2019 17:42	SMPL
K3399-21	MH0BW8	3892	0.8969 -	ISM02.4	PPB	6/19/2019 17:44	SMPL
K3399-22	MH0BW9	3847	0.8865 -	ISM02.4	PPB	6/19/2019 17:46	SMPL
K3391-15DLX2		20671	4.7562 -	ISM02.4	PPB	6/19/2019 17:49	SMPL
CCV05	CCV05	20280	4.6662 -	ISM02.4	PPB	6/19/2019 17:51	SMPL
CCB05	CCB05	64	0.0164 -	ISM02.4	PPB	6/19/2019 17:53	SMPL