

LB103516

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

6/19/2019 16:14

Test: Total CN

ug/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.666	0.0	0.097	
ICB1	1.132	0.0	0.004	
CCV1	243.114	0.0	0.244	
CCB1	1.273	0.0	0.004	
PB120767BL	0.889	0.0	0.004	
PB120767BS	96.075	0.0	0.098	
MIDPB120767	243.320	0.0	0.245	
K3378-02	42.922	0.0	0.045	
K3378-02DUP	43.130	0.0	0.046	
K3378-02MS	129.599	0.0	0.132	
K3378-02MSD	129.008	0.0	0.131	
K3410-02	0.545	0.0	0.003	
CCV2	257.468	0.0	0.259	
CCB2	0.487	0.0	0.003	
K3378-02A	82.336	0.0	0.085	
CCV3	258.430	0.0	0.260	
CCB3	0.449	0.0	0.003	

N	17
Mean	95.579
SD	99.4363
CV%	104.04

Aquakem v. 7.2AQ1

Results from time period:

Wed Jun 19 15:49:39 2019

Wed Jun 19 16:14:18 2019



Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.4309	µg/l	6/19/2019 9:45:10	
5.0PPBCN	A	Total CN	P	5.4643	µg/l	6/19/2019 9:45:11	
10PPBCN	A	Total CN	P	10.4408	µg/l	6/19/2019 9:45:12	
50PPBCN	A	Total CN	P	50.1427	µg/l	6/19/2019 9:45:13	
100PPBCN	A	Total CN	P	98.1555	µg/l	6/19/2019 9:45:14	
250PPBCN	A	Total CN	P	250.0491	µg/l	6/19/2019 9:45:15	
500PPBCN	A	Total CN	P	500.3167	µg/l	6/19/2019 9:45:16	
ICV1	S	Total CN	P	94.6665	µg/l	6/19/2019 15:49:39	
ICB1	S	Total CN	P	1.1317	µg/l	6/19/2019 15:49:40	
CCV1	S	Total CN	P	243.1138	µg/l	6/19/2019 15:49:41	
CCB1	S	Total CN	P	1.2731	µg/l	6/19/2019 15:49:42	
PB120767BL	S	Total CN	P	0.8886	µg/l	6/19/2019 15:49:43	
PB120767BS	S	Total CN	P	96.0754	µg/l	6/19/2019 15:49:44	
MIDPB120767	S	Total CN	P	243.3203	µg/l	6/19/2019 15:49:45	
K3378-02	S	Total CN	P	42.9221	µg/l	6/19/2019 15:49:46	
K3378-02DUP	S	Total CN	P	43.1304	µg/l	6/19/2019 15:49:47	
K3378-02MS	S	Total CN	P	129.599	µg/l	6/19/2019 15:49:48	
K3378-02MSD	S	Total CN	P	129.008	µg/l	6/19/2019 15:49:49	
K3410-02	S	Total CN	P	0.545	µg/l	6/19/2019 15:53:30	
CCV2	S	Total CN	P	257.4684	µg/l	6/19/2019 15:53:31	
CCB2	S	Total CN	P	0.4867	µg/l	6/19/2019 15:53:32	
K3378-02A	S	Total CN	P	82.336	µg/l	6/19/2019 16:14:16	
CCV3	S	Total CN	P	258.4295	µg/l	6/19/2019 16:14:17	
CCB3	S	Total CN	P	0.4493	µg/l	6/19/2019 16:14:18	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

6/19/2019 15:52


Test Total CN

Accepted 6/19/2019 9:48

Factor 1006
Bias 0.003

Coeff. of det. 0.999980

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.4309	0.0000	
2	5.0PPBCN	0.008	5.4643	5.0000	
3	10PPBCN	0.013	10.4408	10.0000	
4	50PPBCN	0.053	50.1427	50.0000	
5	100PPBCN	0.100	98.1555	100.0000	
6	250PPBCN	0.251	250.0491	250.0000	
7	500PPBCN	0.500	500.3167	500.0000	