

**PERCENT SOLID**

Supervisor: apatel
Analyst: JIGNESH
Date: 6/20/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:25
In Date: 06/19/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 06/20/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103517

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3159-05	OR-02-061919-A	1	1.14	9.84	8.83	88.4
K3159-07	OR-02-061919-B	2	1.16	9.79	8.86	89.2
K3447-01	BOTTOM-1	3	1.15	9.96	8.35	81.7
K3447-02	BOTTOM-2	4	1.18	9.98	8.47	82.8
K3447-03	SIDEWALL-1	5	1.14	9.93	8.17	80.0
K3447-04	SIDEWALL-2	6	1.18	9.94	8.78	86.8
K3447-05	SIDEWALL-3	7	1.12	9.65	7.97	80.3
K3447-06	SIDEWALL-4	8	1.17	9.97	8.42	82.4
K3447-07	DISPENSER-1	9	1.15	9.94	8.53	84.0
K3447-08	DISPENSER-2	10	1.12	9.51	8.9	92.7
K3447-09	JUNCTION	11	1.16	9.53	7.44	75.0
K3448-01	M-2 (6-8)	12	1.19	9.7	9.02	92.0
K3448-02	M-2 (0-10)	13	1.16	9.96	8.94	88.4
K3448-03	M-3 (6-8)	14	1.19	9.73	8.62	87.0
K3448-04	M-3 (0-10)	15	1.18	9.68	8.8	89.6
K3450-01	DAM-CONCRETE	16	1.00	2.00	2.00	100.0
K3453-01	FRAC-TANK-MUD	17	1.12	9.73	6.26	59.7
K3453-05	FRAC-TANK-MUD	18	1.12	9.73	6.26	59.7
K3454-03	30272	19	1.00	2.00	2.00	100.0
K3455-01	TP-1-MH	20	1.11	9.59	8.35	85.4
K3455-02	TP-1-MH	21	1.11	9.59	8.35	85.4
K3455-03	TP-2-GIS	22	1.14	9.56	8.66	89.3
K3455-04	TP-2-GIS	23	1.14	9.56	8.66	89.3
K3455-05	TP-3-GIS	24	1.19	9.66	8.98	92.0
K3455-06	TP-3-GIS	25	1.19	9.66	8.98	92.0
K3455-07	TP-4-GIS	26	1.18	9.79	8.65	86.8
K3455-08	TP-4-GIS	27	1.18	9.79	8.65	86.8
K3455-09	TP-5-DUCT-BANK	28	1.14	9.92	9.15	91.2
K3455-10	TP-5-DUCT-BANK	29	1.14	9.92	9.15	91.2

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$