



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103518

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/19/2019	17:25
2	CAL2	1	Water	NA	NA	20.3	7.01	06/19/2019	17:26
3	CAL3	1	Water	NA	NA	20.1	10.03	06/19/2019	17:27
4	ICV	1	Water	NA	NA	20.3	7.00	06/19/2019	17:28
5	CCV1	1	Water	NA	NA	20.2	2.01	06/19/2019	18:00
6	K3159-06	1	Solid	20.02	20	22.4	8.76	06/19/2019	18:01
7	K3159-06DUP	1	Solid	20.03	20	22.5	8.77	06/19/2019	18:02
8	K3159-08	1	Solid	20.04	20	22.1	8.77	06/19/2019	18:04
9	K3448-05	1	Solid	20.02	20	21.6	8.20	06/19/2019	18:05
10	K3448-06	1	Solid	20.03	20	21.4	7.84	06/19/2019	18:06
11	K3453-02	1	Solid	20.04	20	21.0	7.23	06/19/2019	18:07
12	K3455-02	1	Solid	20.03	20	22.1	8.38	06/19/2019	18:10
13	K3455-04	1	Solid	20.04	20	22.2	7.68	06/19/2019	18:11
14	K3455-06	1	Solid	20.01	20	22.1	5.63	06/19/2019	18:12
15	K3455-08	1	Solid	20.04	20	21.8	8.11	06/19/2019	18:15
16	CCV2	1	Water	NA	NA	20.1	12.01	06/19/2019	18:16
17	K3455-10	1	Solid	20.03	20	21.9	9.33	06/19/2019	18:18
18	CCV3	1	Water	NA	NA	20.4	2.02	06/19/2019	18:20

28 103518

WORKLIST(Hardcopy Internal Chain)

Date : 06-19-2019 17:17:08

WorkList ID : 127638

WorkList Name : corrosivity k3161

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2019	Solid	K3159-06	Corrosivity	Cool 4 deg C	PSEG05		OR-02-061919-A	06/19/2019	9045D
06/19/2019	Solid	K3159-08	Corrosivity	Cool 4 deg C	PSEG05		OR-02-061919-B	06/19/2019	9045D
06/18/2019	Solid	K3448-05	Corrosivity	Cool 4 deg C	WSP001	J13	M-2(0-10)	06/18/2019	9045D
06/18/2019	Solid	K3448-06	Corrosivity	Cool 4 deg C	WSP001	J13	M-3(0-10)	06/18/2019	9045D
06/19/2019	Solid	K3453-02	Corrosivity	Cool 4 deg C	AREC01		FRAC-TANK-MUD	06/19/2019	9045D
06/19/2019	Solid	K3455-02	Corrosivity	Cool 4 deg C	PSEG01	K22	TP-1-MH	06/19/2019	9045D
06/19/2019	Solid	K3455-04	Corrosivity	Cool 4 deg C	PSEG01	K22	TP-2-GIS	06/19/2019	9045D
06/19/2019	Solid	K3455-06	Corrosivity	Cool 4 deg C	PSEG01	K22	TP-3-GIS	06/19/2019	9045D
06/19/2019	Solid	K3455-08	Corrosivity	Cool 4 deg C	PSEG01	K22	TP-4-GIS	06/19/2019	9045D
06/19/2019	Solid	K3455-10	Corrosivity	Cool 4 deg C	PSEG01	K22	TP-5-DUCT-BANK	06/19/2019	9045D

Date/Time 06.14.19 17:20
 Received by: Jy
 Relinquished by: Jy

Date/Time 06.14.19 18:30
 Received by: PC
 Relinquished by: Jy