



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103519

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/19/2019	17:25
2	CAL2	1	Water	NA	NA	20.3	7.01	06/19/2019	17:26
3	CAL3	1	Water	NA	NA	20.1	10.03	06/19/2019	17:27
4	ICV	1	Water	NA	NA	20.3	7.00	06/19/2019	17:27
5	CCV1	1	Water	NA	NA	20.3	2.02	06/19/2019	17:30
6	K3451-01	1	Water	NA	NA	21.1	6.95	06/19/2019	17:31
7	K3451-01DUP	1	Water	NA	NA	21.2	6.96	06/19/2019	17:32
8	K3451-02	1	Water	NA	NA	21.2	7.15	06/19/2019	17:33
9	K3451-03	1	Water	NA	NA	21.4	4.92	06/19/2019	17:34
10	K3451-04	1	Water	NA	NA	21.4	4.86	06/19/2019	17:35
11	K3451-05	1	Water	NA	NA	21.5	4.55	06/19/2019	17:37
12	K3451-06	1	Water	NA	NA	21.4	5.84	06/19/2019	17:38
13	K3451-07	1	Water	NA	NA	21.4	5.46	06/19/2019	17:40
14	K3451-08	1	Water	NA	NA	21.6	10.08	06/19/2019	17:42
15	CCV2	1	Water	NA	NA	20.4	12.02	06/19/2019	17:44

LB 103519

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k3451

WorkList ID : 127639

Date : 06-19-2019 17:17:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2019	Water	K3451-01 ✓	pH	Cool 4 deg C	PSEG01	K21	30230	06/19/2019	9040C
06/19/2019	Water	K3451-02 ✓	pH	Cool 4 deg C	PSEG01	K21	30181	06/19/2019	9040C
06/19/2019	Water	K3451-03 ✓	pH	Cool 4 deg C	PSEG01	K21	30191	06/19/2019	9040C
06/19/2019	Water	K3451-04 ✓	pH	Cool 4 deg C	PSEG01	K21	30192	06/19/2019	9040C
06/19/2019	Water	K3451-05 ✓	pH	Cool 4 deg C	PSEG01	K21	30193	06/19/2019	9040C
06/19/2019	Water	K3451-06 ✓	pH	Cool 4 deg C	PSEG01	K21	30231	06/19/2019	9040C
06/19/2019	Water	K3451-07 ✓	pH	Cool 4 deg C	PSEG01	K21	30254	06/19/2019	9040C
06/19/2019	Water	K3451-08 ✓	pH	Cool 4 deg C	PSEG01	K21	30281	06/19/2019	9040C

Date/Time 06-19-19 17:20
 Received by: gn
 Relinquished by: pc

Date/Time 06-19-19 18:25
 Received by: pc
 Relinquished by: gn