



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103532

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	06/20/2019	11:45
2	CAL2	1	Water	NA	NA	21.2	7.00	06/20/2019	11:46
3	CAL3	1	Water	NA	NA	21.2	10.03	06/20/2019	11:47
4	ICV	1	Water	NA	NA	21.1	6.99	06/20/2019	11:48
5	CCV1	1	Water	NA	NA	21.1	2.02	06/20/2019	11:49
6	K3459-01	1	Water	NA	NA	21.7	4.52	06/20/2019	11:50
7	K3459-01DUP	1	Water	NA	NA	21.6	4.53	06/20/2019	11:51
8	K3459-02	1	Water	NA	NA	21.9	4.43	06/20/2019	12:00
9	K3459-03	1	Water	NA	NA	21.2	5.11	06/20/2019	12:02
10	CCV2	1	Water	NA	NA	21.2	12.01	06/20/2019	12:05

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3459

WorkList ID : 127679

Date : 06-20-2019 11:31:27

JP 12103532 JP 12103532

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2019	Water	K3459-01-A	pH	Cool 4 deg C	STER07	J13	CELLS-A	06/19/2019	SM 4500-H B
06/19/2019	Water	K3459-02-A	pH	Cool 4 deg C	STER07	J13	CELLS-B/C	06/19/2019	SM 4500-H B
06/19/2019	Water	K3459-03-A	pH	Cool 4 deg C	STER07	J13	CELLS-D	06/19/2019	SM 4500-H B

Date/Time 06-20-19 11:35
 Received by: Jo
 Relinquished by: Jo

Date/Time 06-20-19 12:35
 Received by: Jo
 Relinquished by: Jo