

LB103549


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51649



A= 0.0000e+000

B= 1.9430e-004

C= 7.4929e-003

Rho= 0.9999083

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	-129	0.000	-129				
0.2	0.200	0.187	-0.013	923	0.0 %	923				
2.5	2.500	2.558	0.058	13128	0.0 %	13128				
5.0	5.000	5.023	0.023	25816	0.0 %	25816				
7.5	7.500	7.406	-0.094	38081	0.0 %	38081				
10.0	10.000	10.043	0.043	51649	0.0 %	51649				

LB103549 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Jun 2019 09:11:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	-129	-	0.0000	7471B	PPB	21 Jun 2019 09:13:14	S	Std
0.2 - 1	50-2	923	-	0.2000	7471B	PPB	21 Jun 2019 09:15:34	S	Std
2.5 - 1	52-5	13128	-	2.5000	7471B	PPB	21 Jun 2019 09:17:52	S	Std
5.0 - 1	55-0	25816	-	5.0000	7471B	PPB	21 Jun 2019 09:20:10	S	Std
7.5 - 1	57-5	38081	-	7.5000	7471B	PPB	21 Jun 2019 09:22:27	S	Std
10.0 - 1	510-0	51649	-	10.0000	7471B	PPB	21 Jun 2019 09:24:44	S	Std
ICV05 - 1	ICV05	22085	4.2985	-	7471B	PPB	21 Jun 2019 09:27:32	U	SMPL
ICB05 - 1	ICB05	-88	-0.0096	-	7471B	PPB	21 Jun 2019 09:29:48	U	SMPL
CCV06 - 1	CCV06	25901	5.0399	-	7471B	PPB	21 Jun 2019 09:32:03	U	SMPL
CCB06 - 1	CCB06	-69	-0.0059	-	7471B	PPB	21 Jun 2019 09:34:19	U	SMPL
CRA - 1	CRA	977	0.1973	-	7471B	PPB	21 Jun 2019 09:36:35	U	SMPL
HighStd - 1	HighStd	51487	10.0112	-	7471B	PPB	21 Jun 2019 09:38:51	U	SMPL
ChkStd - 1	ChkStd	35940	6.9905	-	7471B	PPB	21 Jun 2019 09:41:07	U	SMPL
PB120815BL - 1	PBS	57	0.0186	-	7471B	PPB	21 Jun 2019 09:43:25	U	SMPL
PB120815BS - 1	LCSS	20411	3.9733	-	7471B	PPB	21 Jun 2019 09:45:42	U	SMPL
K3159-05 - 1	OR-02-061919-A	2219	0.4386	-	7471B	PPB	21 Jun 2019 09:47:59	U	SMPL
K3159-07 - 1	OR-02-061919-B	1507	0.3003	-	7471B	PPB	21 Jun 2019 09:50:16	U	SMPL
K3329-01 - 1	WASTE	7779	1.5189	-	7471B	PPB	21 Jun 2019 09:52:34	U	SMPL
K3435-01 - 1	FK-SF-02-004-A	1272	0.2546	-	7471B	PPB	21 Jun 2019 09:54:52	U	SMPL
K3435-03 - 1	FK-SF-02-004-B	966	0.1952	-	7471B	PPB	21 Jun 2019 09:57:07	U	SMPL
CCV07 - 1	CCV07	25868	5.0335	-	7471B	PPB	21 Jun 2019 09:59:23	U	SMPL
CCB07 - 1	CCB07	-72	-0.0065	-	7471B	PPB	21 Jun 2019 10:01:39	U	SMPL
K3438-01 - 1	FK-SF-01-005-A	1198	0.2403	-	7471B	PPB	21 Jun 2019 10:03:55	U	SMPL
K3438-03 - 1	FK-SF-01-005-B	1073	0.2160	-	7471B	PPB	21 Jun 2019 10:06:10	U	SMPL
K3439-01 - 1	FK-0D015-03-006	540	0.1124	-	7471B	PPB	21 Jun 2019 10:08:26	U	SMPL
K3440-01 - 1	FK-SF-03-008-A	1037	0.2090	-	7471B	PPB	21 Jun 2019 10:10:42	U	SMPL
K3440-03 - 1	FK-SF-03-008-B	1516	0.3020	-	7471B	PPB	21 Jun 2019 10:12:59	U	SMPL
K3441-02 - 1	SAMPLE-3	794	0.1618	-	7471B	PPB	21 Jun 2019 10:15:15	U	SMPL
K3441-02DUP - 1	SAMPLE-3DUP	789	0.1608	-	7471B	PPB	21 Jun 2019 10:17:32	U	SMPL
K3441-02MS - 1	SAMPLE-3MS	20966	4.0811	-	7471B	PPB	21 Jun 2019 10:19:49	U	SMPL
K3441-02MSD - 1	SAMPLE-3MSD	21618	4.2078	-	7471B	PPB	21 Jun 2019 10:22:07	U	SMPL
K3448-02 - 1	M-2(0-10)	2950	0.5807	-	7471B	PPB	21 Jun 2019 10:24:24	U	SMPL
CCV08 - 1	CCV08	26268	5.1112	-	7471B	PPB	21 Jun 2019 10:26:41	U	SMPL
CCB08 - 1	CCB08	-80	-0.0081	-	7471B	PPB	21 Jun 2019 10:28:57	U	SMPL
K3448-04 - 1	M-3(0-10)	5579	1.0915	-	7471B	PPB	21 Jun 2019 10:31:13	U	SMPL
K3450-01 - 1	DAM-CONCRETE	2149	0.4250	-	7471B	PPB	21 Jun 2019 10:33:29	U	SMPL
K3453-01 - 1	FRAC-TANK-MUD	84956	16.5141	-	7471B	PPB	21 Jun 2019 10:35:45	U	SMPL
K3455-01 - 1	TP-1-MH	59906	11.6470	-	7471B	PPB	21 Jun 2019 10:38:01	U	SMPL
K3455-03 - 1	TP-2-GIS	22241	4.3288	-	7471B	PPB	21 Jun 2019 10:40:20	U	SMPL
K3455-05 - 1	TP-3-GIS	17502	3.4081	-	7471B	PPB	21 Jun 2019 10:42:37	U	SMPL
K3455-07 - 1	TP-4-GIS	54699	10.6353	-	7471B	PPB	21 Jun 2019 10:44:54	U	SMPL
K3455-09 - 1	TP-5-DUCT-BANK	45400	8.8285	-	7471B	PPB	21 Jun 2019 10:47:11	U	SMPL
K3441-02LX5 - 1		-9	0.0057	-	7471B	PPB	21 Jun 2019 10:49:30	U	SMPL
K3441-02A - 1		21229	4.1322	-	7471B	PPB	21 Jun 2019 10:53:31	U	SMPL
CCV09 - 1	CCV09	26321	5.1215	-	7471B	PPB	21 Jun 2019 10:55:48	U	SMPL
CCB09 - 1	CCB09	-99	-0.0117	-	7471B	PPB	21 Jun 2019 10:58:04	U	SMPL
K3453-01DLX4 - 1		20643	4.0183	-	7471B	PPB	21 Jun 2019 11:00:21	U	SMPL
K3455-01DLX2 - 1		29151	5.6714	-	7471B	PPB	21 Jun 2019 11:02:37	U	SMPL
K3455-07DLX5 - 1		10354	2.0192	-	7471B	PPB	21 Jun 2019 11:04:53	U	SMPL
CCV10 - 1	CCV10	25195	4.9028	-	7471B	PPB	21 Jun 2019 11:07:09	U	SMPL
CCB10 - 1	CCB10	-86	-0.0092	-	7471B	PPB	21 Jun 2019 11:09:25	U	SMPL