

7470A


LB103553

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35330



A= 0.0000e+000

B= 2.7676e-004

C= -3.5500e-002

Rho= 0.9988563

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.030	-0.030	21	0.000	21				
0.2	0.200	0.169	-0.031	738	0.0 %	738				
2.5	2.500	2.520	0.020	9234	0.0 %	9234				
5.0	5.000	4.957	-0.043	18040	0.0 %	18040				
7.5	7.500	7.841	0.341	28460	0.0 %	28460				
10.0	10.000	9.742	-0.258	35330	0.0 %	35330				

LB103553 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Jun 2019 12:34:32

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	21	-	0.0000	7470A	PPB	21 Jun 2019 12:47:53	S	Std
0.2 - 1	50.2	738	-	0.2000	7470A	PPB	21 Jun 2019 12:50:02	S	Std
2.5 - 1	52.5	9234	-	2.5000	7470A	PPB	21 Jun 2019 12:52:11	S	Std
5.0 - 1	55.0	18040	-	5.0000	7470A	PPB	21 Jun 2019 12:54:19	S	Std
7.5 - 1	57.5	28460	-	7.5000	7470A	PPB	21 Jun 2019 12:56:27	S	Std
10.0 - 1	50.0	35330	-	10.0000	7470A	PPB	21 Jun 2019 13:05:03	S	Std
ICV06 - 1	ICV06	14576	3.9986	-	7470A	PPB	21 Jun 2019 13:34:38	U	SMPL
ICB06 - 1	ICB06	-49	-0.0491	-	7470A	PPB	21 Jun 2019 13:36:44	U	SMPL
CCV11 - 1	CCV11	17998	4.9456	-	7470A	PPB	21 Jun 2019 13:38:50	U	SMPL
CCB11 - 1	CCB11	-58	-0.0516	-	7470A	PPB	21 Jun 2019 13:40:56	U	SMPL
CRA - 1	CRA	734	0.1676	-	7470A	PPB	21 Jun 2019 13:43:02	U	SMPL
HighStd - 1	HighStd	34897	9.6226	-	7470A	PPB	21 Jun 2019 13:45:08	U	SMPL
ChkStd - 1	ChkStd	26041	7.1716	-	7470A	PPB	21 Jun 2019 13:47:15	U	SMPL
PB120829BL - 1	PBW	-69	-0.0546	-	7470A	PPB	21 Jun 2019 13:49:22	U	SMPL
PB120829BS - 1	LCSW	14516	3.9820	-	7470A	PPB	21 Jun 2019 13:51:28	U	SMPL
K3433-02 - 1	OILY-WATER	-18	-0.0405	-	7470A	PPB	21 Jun 2019 13:53:36	U	SMPL
K3433-02DUP - 1	OILY-WATERDUP	76	-0.0145	-	7470A	PPB	21 Jun 2019 13:55:44	U	SMPL
K3433-02MS - 1	OILY-WATERMS	14906	4.0899	-	7470A	PPB	21 Jun 2019 13:57:52	U	SMPL
K3433-02MSD - 1	OILY-WATERMSD	14102	3.8674	-	7470A	PPB	21 Jun 2019 14:00:00	U	SMPL
K3441-06 - 1	GWMH-4	-41	-0.0468	-	7470A	PPB	21 Jun 2019 14:02:07	U	SMPL
CCV12 - 1	CCV12	18193	4.9996	-	7470A	PPB	21 Jun 2019 14:04:13	U	SMPL
CCB12 - 1	CCB12	-48	-0.0488	-	7470A	PPB	21 Jun 2019 14:06:20	U	SMPL
PB120837BL - 1	PBW	16	-0.0311	-	7470A	PPB	21 Jun 2019 14:08:26	U	SMPL
PB120837BS - 1	LCSW	14658	4.0213	-	7470A	PPB	21 Jun 2019 14:10:33	U	SMPL
K3441-04 - 1	SAMPLE-3	-46	-0.0482	-	7470A	PPB	21 Jun 2019 14:12:40	U	SMPL
K3441-04DUP - 1	SAMPLE-3DUP	39	-0.0247	-	7470A	PPB	21 Jun 2019 14:14:48	U	SMPL
K3441-04MS - 1	SAMPLE-3MS	14810	4.0633	-	7470A	PPB	21 Jun 2019 14:16:55	U	SMPL
K3441-04MSD - 1	SAMPLE-3MSD	14357	3.9380	-	7470A	PPB	21 Jun 2019 14:19:02	U	SMPL
PB120831BL - 1	PBW	-37	-0.0457	-	7470A	PPB	21 Jun 2019 14:21:11	U	SMPL
PB120831BS - 1	LCSW	14841	4.0719	-	7470A	PPB	21 Jun 2019 14:23:18	U	SMPL
K3159-06 - 1	OR-02-061919-A	614	0.1344	-	7470A	PPB	21 Jun 2019 14:25:27	U	SMPL
K3159-06DUP - 1	OR-02-061919-ADUP	993	0.2393	-	7470A	PPB	21 Jun 2019 14:27:33	U	SMPL
CCV13 - 1	CCV13	18098	4.9733	-	7470A	PPB	21 Jun 2019 14:29:40	U	SMPL
CCB13 - 1	CCB13	-85	-0.0590	-	7470A	PPB	21 Jun 2019 14:31:46	U	SMPL
K3159-06MS - 1	OR-02-061919-AMS	14641	4.0166	-	7470A	PPB	21 Jun 2019 14:33:53	U	SMPL
K3159-06MSD - 1	OR-02-061919-AMSD	14773	4.0531	-	7470A	PPB	21 Jun 2019 14:36:00	U	SMPL
K3159-08 - 1	OR-02-061919-B	379	0.0694	-	7470A	PPB	21 Jun 2019 14:38:06	U	SMPL
K3448-05 - 1	M-2(0-10)	497	0.1021	-	7470A	PPB	21 Jun 2019 14:40:13	U	SMPL
K3448-06 - 1	M-3(0-10)	567	0.1214	-	7470A	PPB	21 Jun 2019 14:42:20	U	SMPL
K3450-02 - 1	DAM-CONCRETE	701	0.1585	-	7470A	PPB	21 Jun 2019 14:44:27	U	SMPL
K3453-02 - 1	FRAC-TANK-MUD	1227	0.3041	-	7470A	PPB	21 Jun 2019 14:46:35	U	SMPL
K3454-03 - 1	30272	-16	-0.0399	-	7470A	PPB	21 Jun 2019 14:48:44	U	SMPL
K3455-02 - 1	TP-1-MH	352	0.0619	-	7470A	PPB	21 Jun 2019 14:50:53	U	SMPL
K3455-04 - 1	TP-2-GIS	354	0.0625	-	7470A	PPB	21 Jun 2019 14:52:59	U	SMPL
CCV14 - 1	CCV14	17833	4.9000	-	7470A	PPB	21 Jun 2019 14:59:47	U	SMPL
CCB14 - 1	CCB14	-75	-0.0563	-	7470A	PPB	21 Jun 2019 15:01:54	U	SMPL
K3455-06 - 1	TP-3-GIS	284	0.0431	-	7470A	PPB	21 Jun 2019 15:04:00	U	SMPL
K3455-08 - 1	TP-4-GIS	319	0.0528	-	7470A	PPB	21 Jun 2019 15:06:07	U	SMPL
K3455-10 - 1	TP-5-DUCT-BANK	299	0.0473	-	7470A	PPB	21 Jun 2019 15:08:13	U	SMPL
PB120832BL - 1	PBW	427	0.0827	-	7470A	PPB	21 Jun 2019 15:10:20	U	SMPL
PB120832BS - 1	LCSW	14929	4.0963	-	7470A	PPB	21 Jun 2019 15:12:27	U	SMPL
K3158-06 - 1	OK-03-062019	36	-0.0255	-	7470A	PPB	21 Jun 2019 15:14:34	U	SMPL
K3158-06DUP - 1	OK-03-062019DUP	20	-0.0300	-	7470A	PPB	21 Jun 2019 15:16:42	U	SMPL
K3158-06MS - 1	OK-03-062019MS	14511	3.9806	-	7470A	PPB	21 Jun 2019 15:18:50	U	SMPL
K3158-06MSD - 1	OK-03-062019MSD	14916	4.0927	-	7470A	PPB	21 Jun 2019 15:20:58	U	SMPL
K3470-04 - 1	SAMPLE-4	82	-0.0128	-	7470A	PPB	21 Jun 2019 15:23:05	U	SMPL
CCV15 - 1	CCV15	17814	4.8947	-	7470A	PPB	21 Jun 2019 15:25:13	U	SMPL
CCB15 - 1	CCB15	-78	-0.0571	-	7470A	PPB	21 Jun 2019 15:27:20	U	SMPL
K3470-08 - 1	SAMPLE-5	190	0.0171	-	7470A	PPB	21 Jun 2019 15:29:27	U	SMPL
K3470-12 - 1	SAMPLE-6	72	-0.0156	-	7470A	PPB	21 Jun 2019 15:31:34	U	SMPL
PB120829TB - 1	PB120741TB	67	-0.0170	-	7470A	PPB	21 Jun 2019 15:33:41	U	SMPL
PB120837TB - 1	PB120739TB	65	-0.0175	-	7470A	PPB	21 Jun 2019 15:35:53	U	SMPL
PB120831TB - 1	PB120773TB	-28	-0.0432	-	7470A	PPB	21 Jun 2019 15:38:05	U	SMPL
PB120832TB - 1	PB120804TB	-28	-0.0432	-	7470A	PPB	21 Jun 2019 15:40:12	U	SMPL
K3433-02LX5 - 1		-63	-0.0529	-	7470A	PPB	21 Jun 2019 15:42:20	U	SMPL
K3433-02A - 1		14643	4.0171	-	7470A	PPB	21 Jun 2019 15:44:29	U	SMPL
K3441-04LX5 - 1		-42	-0.0471	-	7470A	PPB	21 Jun 2019 15:46:37	U	SMPL
K3441-04A - 1		14441	3.9612	-	7470A	PPB	21 Jun 2019 15:48:44	U	SMPL
CCV16 - 1	CCV16	16622	4.5648	-	7470A	PPB	21 Jun 2019 15:50:51	U	SMPL
CCB16 - 1	CCB16	-92	-0.0610	-	7470A	PPB	21 Jun 2019 15:52:58	U	SMPL
K3159-06LX5 - 1		89	-0.0109	-	7470A	PPB	21 Jun 2019 15:55:06	U	SMPL
K3159-06A - 1		14636	4.0152	-	7470A	PPB	21 Jun 2019 15:57:13	U	SMPL
K3158-06LX5 - 1		-93	-0.0612	-	7470A	PPB	21 Jun 2019 15:59:20	U	SMPL
K3158-06A - 1		15008	4.1181	-	7470A	PPB	21 Jun 2019 16:01:27	U	SMPL
CCV17 - 1	CCV17	17441	4.7915	-	7470A	PPB	21 Jun 2019 16:03:35	U	SMPL
CCB17 - 1	CCB17	-44	-0.0477	-	7470A	PPB	21 Jun 2019 16:05:42	U	SMPL