



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB103554

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/21/2019	12:05
2	CAL2	1	Water	NA	NA	20.1	7.00	06/21/2019	12:06
3	CAL3	1	Water	NA	NA	20.1	10.02	06/21/2019	12:07
4	ICV	1	Water	NA	NA	20.1	7.01	06/21/2019	12:08
5	CCV1	1	Water	NA	NA	20.2	2.02	06/21/2019	12:10
6	K2012-01	1	Water	NA	NA	20.3	4.32	06/21/2019	12:11
7	K2012-01DUP	1	Water	NA	NA	20.4	4.33	06/21/2019	12:12
8	K2012-02	1	Water	NA	NA	20.2	4.66	06/21/2019	12:13
9	CCV2	1	Water	NA	NA	20.2	12.01	06/21/2019	12:15

HB 103554

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2012

WorkList ID : 127732

Date : 06-21-2019 11:59:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2019	Water	K2012-01 - 1	pH	Cool 4 deg C	STER05	J22	BOTTLE-1	06/19/2019	SM 4500-H B
06/19/2019	Water	K2012-02 - 1	pH	Cool 4 deg C	STER05	J22	BOTTLE-2	06/19/2019	SM 4500-H B

Date/Time 06-21-19 12:00
Received by: JD
Relinquished by: PC

Date/Time 06-21-19 13:00
Received by: PC
Relinquished by: JD