



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103560

Slope : 97.5

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/21/2019	15:10
2	CAL2	1	Water	NA	NA	20.1	7.00	06/21/2019	15:11
3	CAL3	1	Water	NA	NA	20.1	10.02	06/21/2019	15:12
4	ICV	1	Water	NA	NA	20.1	7.01	06/21/2019	15:15
5	CCV1	1	Water	NA	NA	20.1	2.01	06/21/2019	15:16
6	K3485-01	1	Water	NA	NA	20.7	6.53	06/21/2019	15:18
7	K3485-01DUP	1	Water	NA	NA	20.8	6.54	06/21/2019	15:19
8	CCV2	1	Water	NA	NA	20.1	12.02	06/21/2019	15:20

WorkList Name : PH K3485

WorkList ID : 127738

Date : 06-21-2019 14:47:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/21/2019	Water	K3485-01 -1	pH	Cool 4 deg C	PSEG01	K31	GWMH-7	06/21/2019	9040C

Date/Time 06-21-19 15:00
Received by: Jn
Relinquished by: CP

Date/Time 06-21-19 16:10
Received by: [Signature]
Relinquished by: Jn

LB 103560

WORKLIST(Hardcopy Internal Chain)