



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103562

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/21/2019	15:10
2	CAL2	1	Water	NA	NA	20.1	7.00	06/21/2019	15:11
3	CAL3	1	Water	NA	NA	20.1	10.02	06/21/2019	15:12
4	ICV	1	Water	NA	NA	20.1	7.01	06/21/2019	15:15
5	CCV1	1	Water	NA	NA	20.1	2.02	06/21/2019	15:30
6	K3487-01	1	Solid	20.02	20	21.6	6.73	06/21/2019	15:32
7	K3490-01	1	Solid	20.03	20	20.8	6.33	06/21/2019	15:33
8	K3490-01DUP	1	Solid	20.04	20	20.7	6.34	06/21/2019	15:35
9	CCV2	1	Water	NA	NA	20.2	12.01	06/21/2019	15:36

2B103562

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3487

WorkList ID : 127740

Date : 06-21-2019 14:49:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/20/2019	Solid	K3487-01-1A	pH	Cool 4 deg C	TULL02	J32	FK-PS-01-048	06/20/2019	9045D
06/20/2019	Solid	K3490-01	pH	Cool 4 deg C	TULL02	J32	FK-PS-01-049	06/20/2019	9045D

Date/Time 06-21-19 15:00
Received by: Jn CP
Relinquished by:

Date/Time 06-21-19 14:00
Received by: CP
Relinquished by: Jn