



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103590

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/24/2019	12:02
2	K3485-02		1	8	100.00	>212.0	>212.0	06/24/2019	12:32
3	K3485-02DUP		1	8	100.00	>212.0	>212.0	06/24/2019	12:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 103540

WorkList Name : FP62419

WorkList ID : 127782

Date : 06-24-2019 09:14:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/05/2019	Water	K3485-02	Flash Point	Cool 4 deg C	PSEG01	K31	GWMH-7	06/21/2019	1010A

Date/Time 06/24/19 11:30
 Received by: KS
 Relinquished by: KS

Date/Time 06/24/19 13:00
 Received by: KS
 Relinquished by: KS