



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103592

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/24/2019	17:35
2	CAL2	1	Water	NA	NA	20.2	7.00	06/24/2019	17:36
3	CAL3	1	Water	NA	NA	20.2	10.03	06/24/2019	17:37
4	ICV	1	Water	NA	NA	20.2	7.01	06/24/2019	17:39
5	CCV1	1	Water	NA	NA	20.1	2.02	06/24/2019	17:40
6	K3156-06	1	Solid	20.02	20	22.4	8.40	06/24/2019	17:41
7	K3156-06DUP	1	Solid	20.03	20	22.5	8.41	06/24/2019	17:42
8	K3503-01	1	Solid	20.00	20	20.8	6.48	06/24/2019	17:43
9	K3510-02	1	Solid	20.02	20	21.6	6.31	06/24/2019	17:45
10	K3511-02	1	Solid	20.03	20	22.1	7.96	06/24/2019	17:46
11	K3511-04	1	Solid	20.04	20	21.8	7.24	06/24/2019	17:48
12	K3512-04	1	Solid	20.02	20	21.2	7.01	06/24/2019	17:50
13	K3512-08	1	Solid	20.03	20	21.4	7.81	06/24/2019	18:00
14	CCV2	1	Water	NA	NA	20.2	12.01	06/24/2019	18:01

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K3156 WorkList ID : 127820 Date : 06-24-2019 14:57:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/24/2019	Solid	K3156-06	Corrosivity	Cool 4 deg C	PSEG05		HD-02-062419	06/24/2019	9045D
06/24/2019	Solid	K3503-01	Corrosivity	Cool 4 deg C	PSEG01	K31	OILY-DEBRIS	06/24/2019	9045D
06/24/2019	Solid	K3510-02	Corrosivity	Cool 4 deg C	PSEG05	K33	TR-04-062419	06/24/2019	9045D
06/24/2019	Solid	K3511-02	Corrosivity	Cool 4 deg C	PSEG01	K32	COMP-1	06/24/2019	9045D
06/24/2019	Solid	K3511-04	Corrosivity	Cool 4 deg C	PSEG01	K32	COMP-2	06/24/2019	9045D
06/24/2019	Solid	K3512-04	Corrosivity	Cool 4 deg C	PSEG01	K33	SAMPLE-7	06/24/2019	9045D
06/24/2019	Solid	K3512-08	Corrosivity	Cool 4 deg C	PSEG01	K33	SAMPLE-8	06/24/2019	9045D

Date/Time 06-24-19 16:20
 Received by: jc
 Relinquished by: jc

Date/Time 06-24-19 18:25
 Received by: jc
 Relinquished by: jc