



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103594

Slope : 98.1

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/24/2019	17:35
2	CAL2	1	Water	NA	NA	20.2	7.00	06/24/2019	17:36
3	CAL3	1	Water	NA	NA	20.2	10.03	06/24/2019	17:37
4	ICV	1	Water	NA	NA	20.2	7.01	06/24/2019	17:39
5	CCV1	1	Water	NA	NA	20.2	2.02	06/24/2019	18:10
6	K3511-05	1	Water	NA	NA	21.1	7.52	06/24/2019	18:11
7	K3511-05DUP	1	Water	NA	NA	21.2	7.53	06/24/2019	18:12
8	CCV2	1	Water	NA	NA	20.2	12.01	06/24/2019	18:15

20163594

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3511 WorkList ID : 127822 Date : 06-24-2019 14:58:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/24/2019	Water	K3511-05-18	pH	Cool 4 deg C	PSEG01	K32	N43502	06/24/2019	9040C

Date/Time 06-24-19 13:00
Received by: Jn
Relinquished by: Jn

Date/Time 06-24-19 16:10
Received by: Jn
Relinquished by: Jn