

7471B


LB103603

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53004



A= 0.0000e+000

B= 1.8924e-004

C= -3.8551e-002

Rho= 0.9999315

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	247	0.000	247				
0.2	0.200	0.214	0.014	1335	0.0 %	1335				
2.5	2.500	2.428	-0.072	13036	0.0 %	13036				
5.0	5.000	5.074	0.074	27016	0.0 %	27016				
7.5	7.500	7.484	-0.016	39751	0.0 %	39751				
10.0	10.000	9.992	-0.008	53004	0.0 %	53004				

LB103603 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 25 Jun 2019 08:56:47				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50-0	247	-	0.0000	7471B	PPB	25 Jun 2019 09:04:29	S Std
0.2 - 1	50-2	1335	-	0.2000	7471B	PPB	25 Jun 2019 09:06:48	S Std
2.5 - 1	52-5	13036	-	2.5000	7471B	PPB	25 Jun 2019 09:09:05	S Std
5.0 - 1	55-0	27016	-	5.0000	7471B	PPB	25 Jun 2019 09:11:23	S Std
7.5 - 1	57-5	39751	-	7.5000	7471B	PPB	25 Jun 2019 09:13:40	S Std
10.0 - 1	510-0	53004	-	10.0000	7471B	PPB	25 Jun 2019 09:18:44	S Std
ICV08 - 1	ICV08	22077	4.1392	-	7471B	PPB	25 Jun 2019 09:21:43	U SMPL
ICB08 - 1	ICB08	297	0.0177	-	7471B	PPB	25 Jun 2019 09:24:02	U SMPL
CCV20 - 1	CCV20	27687	5.2008	-	7471B	PPB	25 Jun 2019 09:26:17	U SMPL
CCB20 - 1	CCB20	259	0.0105	-	7471B	PPB	25 Jun 2019 09:28:32	U SMPL
CRA - 1	CRA	1366	0.2199	-	7471B	PPB	25 Jun 2019 09:30:48	U SMPL
HighStd - 1	HighStd	53303	10.0483	-	7471B	PPB	25 Jun 2019 09:33:04	U SMPL
ChkStd - 1	ChkStd	36907	6.9456	-	7471B	PPB	25 Jun 2019 09:35:20	U SMPL
PB120892BL - 1	PBS	-52	-0.0484	-	7471B	PPB	25 Jun 2019 09:37:37	U SMPL
PB120892BS - 1	LCSS	19220	3.5986	-	7471B	PPB	25 Jun 2019 09:39:55	U SMPL
K3156-05 - 1	HD-02-062419	1020	0.1545	-	7471B	PPB	25 Jun 2019 09:42:12	U SMPL
K3156-05DUP - 1	HD-02-062419DUP	1001	0.1509	-	7471B	PPB	25 Jun 2019 09:44:29	U SMPL
K3156-05MS - 1	HD-02-062419MS	21166	3.9668	-	7471B	PPB	25 Jun 2019 09:46:47	U SMPL
K3156-05MSD - 1	HD-02-062419MSD	21091	3.9526	-	7471B	PPB	25 Jun 2019 09:49:04	U SMPL
K3499-01 - 1	B-1(11-12)	3740	0.6692	-	7471B	PPB	25 Jun 2019 09:51:20	U SMPL
CCV21 - 1	CCV21	27753	5.2133	-	7471B	PPB	25 Jun 2019 09:53:36	U SMPL
CCB21 - 1	CCB21	286	0.0156	-	7471B	PPB	25 Jun 2019 09:55:52	U SMPL
K3499-02 - 1	B-2(11-12)	631	0.0809	-	7471B	PPB	25 Jun 2019 09:58:08	U SMPL
K3499-04 - 1	B-4(8-9)	5748	1.0492	-	7471B	PPB	25 Jun 2019 10:00:24	U SMPL
K3499-05 - 1	B-5(10-11)	26449	4.9666	-	7471B	PPB	25 Jun 2019 10:02:40	U SMPL
K3499-06 - 1	B-6(11-12)	778	0.1087	-	7471B	PPB	25 Jun 2019 10:04:56	U SMPL
K3499-07 - 1	B-6(17-18)	214	0.0019	-	7471B	PPB	25 Jun 2019 10:07:14	U SMPL
K3500-01 - 1	B-1(9-10)	1426	0.2313	-	7471B	PPB	25 Jun 2019 10:09:31	U SMPL
K3500-02 - 1	B-2(9-10)	532	0.0621	-	7471B	PPB	25 Jun 2019 10:11:48	U SMPL
K3500-03 - 1	B-3(9-10)	518	0.0595	-	7471B	PPB	25 Jun 2019 10:14:05	U SMPL
K3500-04 - 1	B-4(9-10)	1487	0.2428	-	7471B	PPB	25 Jun 2019 10:16:23	U SMPL
K3500-05 - 1	B-5(9-10)	1592	0.2627	-	7471B	PPB	25 Jun 2019 10:18:39	U SMPL
CCV22 - 1	CCV22	28203	5.2985	-	7471B	PPB	25 Jun 2019 10:20:55	U SMPL
CCB22 - 1	CCB22	289	0.0161	-	7471B	PPB	25 Jun 2019 10:23:11	U SMPL
K3500-06 - 1	B-6(7.5-8)	1676	0.2786	-	7471B	PPB	25 Jun 2019 10:25:29	U SMPL
K3500-07 - 1	B-7(9-10)	977	0.1463	-	7471B	PPB	25 Jun 2019 10:27:45	U SMPL
K3500-08 - 1	B-8(9-10)	293	0.0169	-	7471B	PPB	25 Jun 2019 10:30:01	U SMPL
K3510-01 - 1	TR-04-062419	4168	0.7502	-	7471B	PPB	25 Jun 2019 10:32:17	U SMPL
K3511-01 - 1	COMP-1	5503	1.0028	-	7471B	PPB	25 Jun 2019 10:34:33	U SMPL
K3511-03 - 1	COMP-2	2207	0.3791	-	7471B	PPB	25 Jun 2019 10:36:50	U SMPL
K3512-02 - 1	SAMPLE-7	4663	0.8439	-	7471B	PPB	25 Jun 2019 10:39:07	U SMPL
K3512-06 - 1	SAMPLE-8	2246	0.3865	-	7471B	PPB	25 Jun 2019 10:41:25	U SMPL
PB120893BL - 1	PBS	50	-0.0291	-	7471B	PPB	25 Jun 2019 10:43:43	U SMPL
PB120893BS - 1	LCSS	19585	3.6676	-	7471B	PPB	25 Jun 2019 10:45:59	U SMPL
CCV23 - 1	CCV23	28423	5.3401	-	7471B	PPB	25 Jun 2019 10:48:14	U SMPL
CCB23 - 1	CCB23	168	-0.0068	-	7471B	PPB	25 Jun 2019 10:50:31	U SMPL
K2459-01 - 1	FK-BPM-04-002	1001	0.1509	-	7471B	PPB	25 Jun 2019 10:52:47	U SMPL
K2459-01DUP - 1	FK-BPM-04-002DUP	998	0.1503	-	7471B	PPB	25 Jun 2019 10:55:03	U SMPL
K2459-01MS - 1	FK-BPM-04-002MS	20514	3.8434	-	7471B	PPB	25 Jun 2019 10:57:20	U SMPL
K2459-01MSD - 1	FK-BPM-04-002MSD	21000	3.9354	-	7471B	PPB	25 Jun 2019 10:59:35	U SMPL
K3152-01 - 1	FK-BPM-04-003	1023	0.1550	-	7471B	PPB	25 Jun 2019 11:01:52	U SMPL
K3155-02 - 1	BAT-B11	227	0.0044	-	7471B	PPB	25 Jun 2019 11:04:09	U SMPL
K3155-04 - 1	BAT-B06	13406	2.4984	-	7471B	PPB	25 Jun 2019 11:06:26	U SMPL
K3158-05 - 1	OK-03-062019	1336	0.2143	-	7471B	PPB	25 Jun 2019 11:08:43	U SMPL
K3164-01 - 1	BAT-B04	17389	3.2521	-	7471B	PPB	25 Jun 2019 11:11:01	U SMPL
K3164-03 - 1	BAT-B05	2324	0.4012	-	7471B	PPB	25 Jun 2019 11:13:17	U SMPL
CCV24 - 1	CCV24	28707	5.3939	-	7471B	PPB	25 Jun 2019 11:15:33	U SMPL
CCB24 - 1	CCB24	276	0.0137	-	7471B	PPB	25 Jun 2019 11:17:50	U SMPL
K3164-05 - 1	BAT-B07	4897	0.8881	-	7471B	PPB	25 Jun 2019 11:20:07	U SMPL
K3164-07 - 1	BAT-B09	469	0.0502	-	7471B	PPB	25 Jun 2019 11:22:23	U SMPL
K3164-09 - 1	BAT-B08	1699	0.2830	-	7471B	PPB	25 Jun 2019 11:24:39	U SMPL
K3470-02 - 1	SAMPLE-4	1739	0.2905	-	7471B	PPB	25 Jun 2019 11:26:56	U SMPL
K3470-06 - 1	SAMPLE-5	770	0.1072	-	7471B	PPB	25 Jun 2019 11:29:12	U SMPL
K3470-10 - 1	SAMPLE-6	394	0.0360	-	7471B	PPB	25 Jun 2019 11:31:29	U SMPL
K3471-01 - 1	WASTE	11138	2.0692	-	7471B	PPB	25 Jun 2019 11:33:45	U SMPL
K3482-01 - 1	TP-1	83950	15.8479	-	7471B	PPB	25 Jun 2019 11:36:03	U SMPL
K3482-03 - 1	TP-2	36970	6.9575	-	7471B	PPB	25 Jun 2019 11:38:20	U SMPL
K3482-05 - 1	TP-3	631	0.0809	-	7471B	PPB	25 Jun 2019 11:40:39	U SMPL
CCV25 - 1	CCV25	27780	5.2184	-	7471B	PPB	25 Jun 2019 11:42:57	U SMPL
CCB25 - 1	CCB25	14	-0.0359	-	7471B	PPB	25 Jun 2019 11:45:14	U SMPL
K3482-07 - 1	TP-4	22777	4.2717	-	7471B	PPB	25 Jun 2019 11:47:31	U SMPL
K3487-01 - 1	FK-PS-01-048	2345	0.4052	-	7471B	PPB	25 Jun 2019 11:49:47	U SMPL
K3156-05LX5 - 1		456	0.0477	-	7471B	PPB	25 Jun 2019 11:52:05	U SMPL
K3156-05A - 1		21678	4.0637	-	7471B	PPB	25 Jun 2019 11:54:21	U SMPL
K2459-01LX5 - 1		427	0.0423	-	7471B	PPB	25 Jun 2019 11:56:37	U SMPL
K2459-01A - 1		20716	3.8817	-	7471B	PPB	25 Jun 2019 11:58:55	U SMPL
K3482-01DLX5 - 1		16128	3.0135	-	7471B	PPB	25 Jun 2019 12:01:12	U SMPL
CCV26 - 1	CCV26	27416	5.1496	-	7471B	PPB	25 Jun 2019 12:03:29	U SMPL
CCB26 - 1	CCB26	-104	-0.0582	-	7471B	PPB	25 Jun 2019 12:05:47	U SMPL