



TOTAL Dissolved Solids - SM2540C

Run Number: LB103605

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/25/2019

BalanceID: WC SC-3

Filter ID: 14466061

TEMP1 IN:	<u>103</u>	<u>06/24/2019</u>	<u>15:00</u>	TEMP OUT:	<u>103</u>	<u>06/25/2019</u>	<u>07:30</u>	OvenID1:	<u>WC Oven-2</u>	Thermo ID1:	<u>WET-OVEN#2</u>
TEMP2 IN:	<u>182</u>	<u>06/25/2019</u>	<u>07:32</u>	TEMP OUT:	<u>182</u>	<u>06/25/2019</u>	<u>09:00</u>	OvenID2:	<u>WC-OVEN-3</u>	Thermo ID2:	<u>WET-OVEN#3</u>
TEMP3 IN:	<u>182</u>	<u>06/25/2019</u>	<u>10:30</u>	TEMP OUT:	<u>182</u>	<u>06/25/2019</u>	<u>12:05</u>	OvenID3:	<u>WC-OVEN-3</u>	Thermo ID3:	<u>WET-OVEN#3</u>

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Final Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Weight (g)	Result mg/L
1	LB103605BL	LB103605BL	89.5682	89.5682	100	89.5682	89.5682	89.5682	0	0
2	LB103605BS	LB103605BS	90.1526	90.1526	100	90.163	90.163	90.163	0.0104	104
3	K3484-01	WASTE-SLUDGE	75.5648	75.5648	100	75.667	75.667	75.667	0.1022	1022
4	K3484-01DUP	WASTE-SLUDGE DUP	92.8199	92.8199	100	92.9199	92.9199	92.9199	0.1	1000

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = 2nd Empty Dish + Sample weight after 1.5 hr drying @180 (±2) °C (g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TDS062419

WorkList ID : 127757

Date : 06-24-2019 08:30:39

13/03605

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/27/2019	Water	K3484-01	TDS	Cool 4 deg C	HOLL01	J31	WASTE-SLUDGE	06/20/2019	SM2540 C

Date/Time 06/24/19 14:00
 Received by: AS
 Relinquished by: CS

Date/Time 06/24/19 16:00
 Received by: CS
 Relinquished by: AS