



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103614

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 8                 | 27.00      | 80.6      | 80.6            | 06/25/2019 | 12:58     |
| 2   | K3511-06    |               | 1  | 7                 | 100.00     | >212.0    | >212.0          | 06/25/2019 | 13:20     |
| 3   | K3511-06DUP |               | 1  | 7                 | 100.00     | >212.0    | >212.0          | 06/25/2019 | 13:50     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

LB 103614

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP62519

WorkList ID : 127861

Date : 06-25-2019 09:54:48

| Due Date   | Matrix | Sample   | Test          | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|---------------|--------------|----------|------------------|-----------------|--------------|--------|
| 06/27/2019 | Water  | K3511-06 | C Flash Point | Cool 4 deg C | PSEG01   | K32              | N43502          | 06/24/2019   | 1010A  |

Date/Time 06/25/19 12:25  
Received by: KS  
Relinquished by: CP

Date/Time 06/25/19 13:55  
Received by: CP  
Relinquished by: KS