


LB103617

ISM02.4

INSTRUMENT ID : CU1

Linear



A= 0.0000e+000

B= 2.5075e-004

C= -4.3680e-002

Rho= 0.9998931

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.036	-0.036	29	0.000	29					—
0.05	0.050										
0.20	0.200	0.162	-0.038	819	0.0 %	819					-19
2.50	2.500	2.551	0.051	10349	0.0 %	10349					2
5.00	5.000	5.096	0.096	20498	0.0 %	20498					2
7.50	7.500	7.456	-0.044	29909	0.0 %	29909					-1
10.0	10.000	9.971	-0.029	39938	0.0 %	39938					0

LB103617
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	29	-		0 ISM02.4	PPB	6/25/2019 15:49	Std
	0.2 S0.2	819	-		0.2 ISM02.4	PPB	6/25/2019 15:51	Std
	2.5 S2.5	10349	-		2.5 ISM02.4	PPB	6/25/2019 15:53	Std
	5 S5.0	20498	-		5 ISM02.4	PPB	6/25/2019 15:55	Std
	7.5 S7.5	29909	-		7.5 ISM02.4	PPB	6/25/2019 15:58	Std
	10 S10.0	39938	-		10 ISM02.4	PPB	6/25/2019 16:00	Std
ICV	ICV	15510	3.8455 -		ISM02.4	PPB	6/25/2019 16:04	SMPL
ICB	ICB	-103	-0.0695 -		ISM02.4	PPB	6/25/2019 16:06	SMPL
CCV032	CCV032	20812	5.175 -		ISM02.4	PPB	6/25/2019 16:09	SMPL
CCB032	CCB032	-90	-0.0662 -		ISM02.4	PPB	6/25/2019 16:11	SMPL
PB120911BL	PBW911	-19	-0.0484 -		ISM02.4	PPB	6/25/2019 16:13	SMPL
K3405-01	MH0CC4	-34	-0.0522 -		ISM02.4	PPB	6/25/2019 16:15	SMPL
K3405-02	MH0CC6	-125	-0.075 -		ISM02.4	PPB	6/25/2019 16:18	SMPL
K3405-03	MH0CD3	-22	-0.0492 -		ISM02.4	PPB	6/25/2019 16:20	SMPL
K3405-04	MH0CE2	4610	1.1123 -		ISM02.4	PPB	6/25/2019 16:22	SMPL
K3405-05	MH0CF5	2239	0.5178 -		ISM02.4	PPB	6/25/2019 16:24	SMPL
K3405-06	MH0CF6	-140	-0.0788 -		ISM02.4	PPB	6/25/2019 16:27	SMPL
K3405-07	MH0BU8	-126	-0.0753 -		ISM02.4	PPB	6/25/2019 16:29	SMPL
K3405-08	MH0BU9	115	-0.0148 -		ISM02.4	PPB	6/25/2019 16:31	SMPL
K3405-09	MH0BW0	-46	-0.0552 -		ISM02.4	PPB	6/25/2019 16:34	SMPL
K3405-10	MH0BW6	-149	-0.081 -		ISM02.4	PPB	6/25/2019 16:36	SMPL
K3405-11	MH0BY1	-137	-0.078 -		ISM02.4	PPB	6/25/2019 16:38	SMPL
K3405-12	MH0CA8	-118	-0.0733 -		ISM02.4	PPB	6/25/2019 16:40	SMPL
K3405-13	MH0CB3	-62	-0.0592 -		ISM02.4	PPB	6/25/2019 16:43	SMPL
K3405-14	MH0CB6	5423	1.3162 -		ISM02.4	PPB	6/25/2019 16:45	SMPL
K3405-15	MH0CB7	723	0.1376 -		ISM02.4	PPB	6/25/2019 16:47	SMPL
K3405-16	MH0CB9	-140	-0.0788 -		ISM02.4	PPB	6/25/2019 16:50	SMPL
K3405-17	MH0BZ1	-134	-0.0773 -		ISM02.4	PPB	6/25/2019 16:52	SMPL
K3405-18	MH0BV5	-99	-0.0685 -		ISM02.4	PPB	6/25/2019 16:54	SMPL
K3405-19	MH0BV5D	-96	-0.0678 -		ISM02.4	PPB	6/25/2019 16:56	SMPL
CCV033	CCV033	20660	5.1369 -		ISM02.4	PPB	6/25/2019 16:59	SMPL
CCB033	CCB033	-57	-0.058 -		ISM02.4	PPB	6/25/2019 17:01	SMPL
K3405-20	MH0BV5S	4438	1.0692 -		ISM02.4	PPB	6/25/2019 17:03	SMPL
PB120911TB	LEB911	-120	-0.0738 -		ISM02.4	PPB	6/25/2019 17:06	SMPL
CCV34	CCV34	19154	4.7593 -		ISM02.4	PPB	6/25/2019 17:08	SMPL
CCB34	CCB34	-65	-0.06 -		ISM02.4	PPB	6/25/2019 17:10	SMPL