



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103623

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/25/2019	18:00
2	CAL2	1	Water	NA	NA	20.2	7.00	06/25/2019	18:01
3	CAL3	1	Water	NA	NA	20.2	10.03	06/25/2019	18:02
4	ICV	1	Water	NA	NA	20.1	7.01	06/25/2019	18:04
5	CCV1	1	Water	NA	NA	20.1	2.02	06/25/2019	18:05
6	K3534-01	1	Solid	20.02	20	22.0	7.00	06/25/2019	18:06
7	K3534-01DUP	1	Solid	20.03	20	22.1	7.01	06/25/2019	18:07
8	CCV2	1	Water	NA	NA	20.2	12.01	06/25/2019	18:09

WORKLIST(Hardcopy Internal Chain)

1B103623

WorkList Name : PH K3534

WorkList ID : 127876

Date : 06-25-2019 17:47:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/25/2019	Solid	K3534-01	pH	Cool 4 deg C	TULL02	J43	FK-PS-01-050	06/25/2019	9045D

Date/Time 06.25.19 17:50
 Received by: Jn
 Relinquished by: PC

Date/Time 06.25.19 18:25
 Received by: PC
 Relinquished by: Jn