

LB/03630

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

6/26/2019 8:16

Test: Phenolics-

Sample Id	Result <i>mg/d</i>	Dil. 1 +	Response	Errors
ICV1	0.999	0.0	0.123	
ICB1	-0.005	0.0	0.014	
CCV1	1.041	0.0	0.127	
CCB1	-0.001	0.0	0.015	
PB120882BL	0.001	0.0	0.015	
PB120882BS	1.008	0.0	0.124	
K3484-01	0.001	0.0	0.015	
K3484-01DUP	0.003	0.0	0.015	
K3484-01MS	1.124	0.0	0.136	
K3484-01MSD	1.144	0.0	0.138	
CCV2	1.014	0.0	0.124	
CCB2	-0.002	0.0	0.015	

N	12
Mean	0.527
SD	0.5529
CV%	104.91

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Jun 26 07:41:59 2019

Wed Jun 26 08:11:13 2019

Sample Id	Sam/Ctr/c	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0481	mg/l	6/26/2019 7:41:59	
0.1PPM	A	Phenolics-	P	0.0969	mg/l	6/26/2019 7:42:00	
0.25PPM	A	Phenolics-	P	0.2498	mg/l	6/26/2019 7:42:01	
0.50PPM	A	Phenolics-	P	0.5022	mg/l	6/26/2019 7:42:02	
1.0PPM	A	Phenolics-	P	1.0066	mg/l	6/26/2019 7:42:03	
2.0PPM	A	Phenolics-	P	1.9964	mg/l	6/26/2019 7:42:04	
ICV1	S	Phenolics-	P	0.9989	mg/l	6/26/2019 8:06:15	
ICB1	S	Phenolics-	P	-0.0055	mg/l	6/26/2019 8:06:16	
CCV1	S	Phenolics-	P	1.041	mg/l	6/26/2019 8:06:17	
CCB1	S	Phenolics-	P	-0.0009	mg/l	6/26/2019 8:06:18	
PB120882BL	S	Phenolics-	P	0.0007	mg/l	6/26/2019 8:06:19	
PB120882BS	S	Phenolics-	P	1.0077	mg/l	6/26/2019 8:06:20	
K3484-01	S	Phenolics-	P	0.0007	mg/l	6/26/2019 8:06:21	
K3484-01DUP	S	Phenolics-	P	0.0027	mg/l	6/26/2019 8:06:22	
K3484-01MS	S	Phenolics-	P	1.1236	mg/l	6/26/2019 8:06:25	
K3484-01MSD	S	Phenolics-	P	1.1443	mg/l	6/26/2019 8:06:26	
CCV2	S	Phenolics-	P	1.0141	mg/l	6/26/2019 8:11:10	
CCB2	S	Phenolics-	P	-0.0021	mg/l	6/26/2019 8:11:11	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

6/26/2019 7:43


Test Phenolics-

Accepted 6/26/2019 7:43

Factor 9.26
Bias 0.015

Coeff. of det. 0.999973

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.020	0.0481	0.0500	
2	Phenol-2	0.025	0.0969	0.1000	
3	Phenol-2	0.042	0.2498	0.2500	
4	Phenol-2	0.069	0.5022	0.5000	
5	Phenol-2	0.124	1.0066	1.0000	
6	Phenol-2	0.230	1.9964	2.0000	