


LB103634

7470A

INSTRUMENT ID : CV1

Linear ▼



A= 0.0000e+000

B= 2.4997e-004

C= 5.0490e-003

Rho= 0.9999577

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	-51	0.000	-51				
0.2	0.200	0.198	-0.002	770	0.0 %	770				
2.5	2.500	2.491	-0.009	9945	0.0 %	9945				
5.0	5.000	5.002	0.002	19991	0.0 %	19991				
7.5	7.500	7.566	0.066	30248	0.0 %	30248				
10.0	10.000	9.951	-0.049	39789	0.0 %	39789				

LB103634 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 26 Jun 2019 09:51:46				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50.0	-51	-	0.0000	7470A	PPB	26 Jun 2019 09:59:40	S Std
0.2 - 1	50.2	770	-	0.2000	7470A	PPB	26 Jun 2019 10:01:50	S Std
2.5 - 1	52.5	9945	-	2.5000	7470A	PPB	26 Jun 2019 10:03:59	S Std
5.0 - 1	55.0	19991	-	5.0000	7470A	PPB	26 Jun 2019 10:06:07	S Std
7.5 - 1	57.5	30248	-	7.5000	7470A	PPB	26 Jun 2019 10:08:15	S Std
10.0 - 1	510.0	39789	-	10.0000	7470A	PPB	26 Jun 2019 10:17:35	S Std
ICV11 - 1	ICV11	15376	3.8485	-	7470A	PPB	26 Jun 2019 10:20:15	U SMPL
ICB11 - 1	ICB11	-88	-0.0169	-	7470A	PPB	26 Jun 2019 10:22:22	U SMPL
CCV35 - 1	CCV35	19788	4.9514	-	7470A	PPB	26 Jun 2019 10:24:28	U SMPL
CCB35 - 1	CCB35	-104	-0.0209	-	7470A	PPB	26 Jun 2019 10:26:34	U SMPL
CRA - 1	CRA	768	0.1970	-	7470A	PPB	26 Jun 2019 10:28:40	U SMPL
HighStd - 1	HighStd	39702	9.9293	-	7470A	PPB	26 Jun 2019 10:30:47	U SMPL
ChkStd - 1	ChkStd	27807	6.9559	-	7470A	PPB	26 Jun 2019 10:32:53	U SMPL
PB120938BL - 1	PBW	-111	-0.0227	-	7470A	PPB	26 Jun 2019 10:35:03	U SMPL
PB120938BS - 1	LCSW	15378	3.8490	-	7470A	PPB	26 Jun 2019 10:37:12	U SMPL
K3156-06 - 1	HD-02-062419	-124	-0.0259	-	7470A	PPB	26 Jun 2019 10:39:20	U SMPL
K3156-06DUP - 1	HD-02-062419DUP	-82	-0.0154	-	7470A	PPB	26 Jun 2019 10:41:28	U SMPL
K3156-06MS - 1	HD-02-062419MS	16242	4.0650	-	7470A	PPB	26 Jun 2019 10:43:36	U SMPL
K3156-06MSD - 1	HD-02-062419MSD	15792	3.9525	-	7470A	PPB	26 Jun 2019 10:45:43	U SMPL
K3475-01 - 1	SOILS-1	-84	-0.0159	-	7470A	PPB	26 Jun 2019 10:47:51	U SMPL
CCV36 - 1	CCV36	20177	5.0486	-	7470A	PPB	26 Jun 2019 10:49:58	U SMPL
CCB36 - 1	CCB36	-81	-0.0152	-	7470A	PPB	26 Jun 2019 10:52:04	U SMPL
K3475-02 - 1	SOILS-2	-58	-0.0094	-	7470A	PPB	26 Jun 2019 10:54:11	U SMPL
K3475-03 - 1	SOILS-3	-75	-0.0137	-	7470A	PPB	26 Jun 2019 10:56:17	U SMPL
K3475-04 - 1	SOILS-4	-92	-0.0179	-	7470A	PPB	26 Jun 2019 10:58:24	U SMPL
K3503-01 - 1	OILY-DEBRIS	-77	-0.0142	-	7470A	PPB	26 Jun 2019 11:00:31	U SMPL
K3510-02 - 1	TR-04-062419	-47	-0.0067	-	7470A	PPB	26 Jun 2019 11:02:38	U SMPL
K3511-02 - 1	COMP-1	-59	-0.0097	-	7470A	PPB	26 Jun 2019 11:04:45	U SMPL
K3511-04 - 1	COMP-2	-12	0.0020	-	7470A	PPB	26 Jun 2019 11:06:53	U SMPL
K3512-04 - 1	SAMPLE-7	-36	-0.0039	-	7470A	PPB	26 Jun 2019 11:09:03	U SMPL
K3512-08 - 1	SAMPLE-8	-29	-0.0022	-	7470A	PPB	26 Jun 2019 11:11:12	U SMPL
PB120939BL - 1	PBW	-31	-0.0027	-	7470A	PPB	26 Jun 2019 11:13:18	U SMPL
CCV37 - 1	CCV37	19857	4.9686	-	7470A	PPB	26 Jun 2019 11:15:25	U SMPL
CCB37 - 1	CCB37	-96	-0.0189	-	7470A	PPB	26 Jun 2019 11:17:32	U SMPL
PB120939BS - 1	LCSW	15533	3.8878	-	7470A	PPB	26 Jun 2019 11:19:39	U SMPL
K3511-06 - 1	N43502	-69	-0.0122	-	7470A	PPB	26 Jun 2019 11:21:45	U SMPL
K3511-06DUP - 1	N43502DUP	-56	-0.0089	-	7470A	PPB	26 Jun 2019 11:23:52	U SMPL
K3511-06MS - 1	N43502MS	17074	4.2730	-	7470A	PPB	26 Jun 2019 11:25:59	U SMPL
K3511-06MSD - 1	N43502MSD	15538	3.8890	-	7470A	PPB	26 Jun 2019 11:28:07	U SMPL
PB120940BL - 1	PBW	-79	-0.0147	-	7470A	PPB	26 Jun 2019 11:30:15	U SMPL
PB120940BS - 1	LCSW	15694	3.9280	-	7470A	PPB	26 Jun 2019 11:32:23	U SMPL
K3520-01 - 1	MW-8R-062019	1291	0.3278	-	7470A	PPB	26 Jun 2019 11:41:26	U SMPL
K3520-01DUP - 1	MW-8R-062019DUP	1281	0.3253	-	7470A	PPB	26 Jun 2019 11:43:34	U SMPL
K3520-01MS - 1	MW-8R-062019MS	16123	4.0353	-	7470A	PPB	26 Jun 2019 11:45:41	U SMPL
CCV38 - 1	CCV38	20011	5.0071	-	7470A	PPB	26 Jun 2019 11:47:48	U SMPL
CCB38 - 1	CCB38	-110	-0.0224	-	7470A	PPB	26 Jun 2019 11:49:55	U SMPL
K3520-01MSD - 1	MW-8R-062019MSD	16150	4.0420	-	7470A	PPB	26 Jun 2019 11:52:02	U SMPL
K3520-02 - 1	SW-3-062019	-9	0.0028	-	7470A	PPB	26 Jun 2019 11:54:08	U SMPL
K3520-04 - 1	SW-2-062019	-157	-0.0342	-	7470A	PPB	26 Jun 2019 11:56:15	U SMPL
K3520-06 - 1	MW-9R-062419	1022	0.2605	-	7470A	PPB	26 Jun 2019 11:58:22	U SMPL
PB120938TB - 1	PB120868TB	-20	0.0000	-	7470A	PPB	26 Jun 2019 12:00:30	U SMPL
PB120939TB - 1	PB120884TB	-27	-0.0017	-	7470A	PPB	26 Jun 2019 12:02:37	U SMPL
K3156-06LX5 - 1		-111	-0.0227	-	7470A	PPB	26 Jun 2019 12:04:44	U SMPL
K3156-06A - 1		16203	4.0553	-	7470A	PPB	26 Jun 2019 12:06:52	U SMPL
K3511-06LX5 - 1		-117	-0.0242	-	7470A	PPB	26 Jun 2019 12:09:01	U SMPL
K3511-06A - 1		16783	4.2002	-	7470A	PPB	26 Jun 2019 12:11:08	U SMPL
CCV39 - 1	CCV39	20324	5.0854	-	7470A	PPB	26 Jun 2019 12:13:15	U SMPL
CCB39 - 1	CCB39	-66	-0.0114	-	7470A	PPB	26 Jun 2019 12:15:22	U SMPL
K3520-01LX5 - 1		164	0.0460	-	7470A	PPB	26 Jun 2019 12:17:31	U SMPL
K3520-01A - 1		16237	4.0638	-	7470A	PPB	26 Jun 2019 12:19:38	U SMPL
CCV40 - 1	CCV40	19387	4.8512	-	7470A	PPB	26 Jun 2019 12:21:45	U SMPL
CCB40 - 1	CCB40	-118	-0.0244	-	7470A	PPB	26 Jun 2019 12:23:52	U SMPL