


LB103638

ISM02.4

INSTRUMENT ID: CU1



Linear

A= 0.0000e+000

B= 2.2311e-004

C= -6.1379e-002

Rho= 0.9996341

Accept=Accepted

Slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.028	-0.028	149	0.000	149					—
0.05	0.050										
0.20	0.200	0.202	0.002	1181	0.0 %	1181					1
2.50	2.500	2.466	-0.034	11330	0.0 %	11330					-1
5.00	5.000	5.188	0.188	23530	0.0 %	23530					4
7.50	7.500	7.352	-0.148	33226	0.0 %	33226					-2
10.0	10.000	10.019	0.019	45182	0.0 %	45182					0

PS 06-26-19

LB103638
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	149	-		0 ISM02.4	PPB	6/26/2019 13:32	Std
	0.2 S0.2	1181	-		0.2 ISM02.4	PPB	6/26/2019 13:35	Std
	2.5 S2.5	11330	-		2.5 ISM02.4	PPB	6/26/2019 13:37	Std
	5 S5.0	23530	-		5 ISM02.4	PPB	6/26/2019 13:39	Std
	7.5 S7.5	33226	-		7.5 ISM02.4	PPB	6/26/2019 13:42	Std
	10 S10.0	45182	-		10 ISM02.4	PPB	6/26/2019 13:44	Std
ICV	ICV	18671	4.1044 -		ISM02.4	PPB	6/26/2019 13:47	SMPL
ICB	ICB	88	-0.0417 -		ISM02.4	PPB	6/26/2019 13:49	SMPL
CCV041	CCV041	23113	5.0954 -		ISM02.4	PPB	6/26/2019 13:51	SMPL
CCB041	CCB041	68	-0.0462 -		ISM02.4	PPB	6/26/2019 13:54	SMPL
PB120934BL	PBS934	115	-0.0357 -		ISM02.4	PPB	6/26/2019 13:56	SMPL
K3401-01	MH0CG2	18726	4.1166 -		ISM02.4	PPB	6/26/2019 13:58	SMPL
K3401-02	MH0CG2D	17623	3.8705 -		ISM02.4	PPB	6/26/2019 14:00	SMPL
K3401-03	MH0CG2S	24795	5.4707 -		ISM02.4	PPB	6/26/2019 14:03	SMPL
K3401-04	MH0BX0	7186	1.5419 -		ISM02.4	PPB	6/26/2019 14:05	SMPL
K3401-05	MH0BX1	543	0.0598 -		ISM02.4	PPB	6/26/2019 14:07	SMPL
K3401-07	MH0BX9	1598	0.2952 -		ISM02.4	PPB	6/26/2019 14:10	SMPL
K3401-08	MH0BY0	2740	0.5499 -		ISM02.4	PPB	6/26/2019 14:12	SMPL
K3401-09	MH0BY1	2396	0.4732 -		ISM02.4	PPB	6/26/2019 14:14	SMPL
K3401-10	MH0BY2	35747	7.9142 -		ISM02.4	PPB	6/26/2019 14:16	SMPL
K3401-11	MH0BY3	1720	0.3224 -		ISM02.4	PPB	6/26/2019 14:19	SMPL
K3401-12	MH0BY4	582	0.0685 -		ISM02.4	PPB	6/26/2019 14:21	SMPL
K3401-13	MH0BY5	781	0.1129 -		ISM02.4	PPB	6/26/2019 14:23	SMPL
K3401-14	MH0BY6	894	0.1381 -		ISM02.4	PPB	6/26/2019 14:26	SMPL
K3401-15	MH0BY7	10998	2.3924 -		ISM02.4	PPB	6/26/2019 14:28	SMPL
K3401-16	MH0BY8	4889	1.0294 -		ISM02.4	PPB	6/26/2019 14:30	SMPL
K3401-17	MH0BY9	4528	0.9489 -		ISM02.4	PPB	6/26/2019 14:32	SMPL
K3401-18	MH0BZ0	12475	2.722 -		ISM02.4	PPB	6/26/2019 14:35	SMPL
K3401-19	MH0BZ2	3338	0.6834 -		ISM02.4	PPB	6/26/2019 14:37	SMPL
K3401-20	MH0BZ3	503	0.0508 -		ISM02.4	PPB	6/26/2019 14:39	SMPL
CCV042	CCV042	22852	5.0372 -		ISM02.4	PPB	6/26/2019 14:41	SMPL
CCB042	CCB042	97	-0.0397 -		ISM02.4	PPB	6/26/2019 14:44	SMPL
K3401-21	MH0BZ4	4533	0.95 -		ISM02.4	PPB	6/26/2019 14:46	SMPL
K3401-22	MH0BZ5	1775	0.3346 -		ISM02.4	PPB	6/26/2019 14:48	SMPL
PB120935BL	PBS935	98	-0.0395 -		ISM02.4	PPB	6/26/2019 14:51	SMPL
K3403-01	MH0BZ6	9272	2.0073 -		ISM02.4	PPB	6/26/2019 14:53	SMPL
K3403-02	MH0BZ7	3904	0.8097 -		ISM02.4	PPB	6/26/2019 14:55	SMPL
K3403-03	MH0BZ8	4264	0.89 -		ISM02.4	PPB	6/26/2019 14:57	SMPL
K3403-04	MH0BZ8D	4078	0.8485 -		ISM02.4	PPB	6/26/2019 15:00	SMPL
K3403-05	MH0BZ8S	16883	3.7054 -		ISM02.4	PPB	6/26/2019 15:02	SMPL
K3403-06	MH0BZ9	3167	0.6452 -		ISM02.4	PPB	6/26/2019 15:04	SMPL
K3403-07	MH0CA0	2344	0.4616 -		ISM02.4	PPB	6/26/2019 15:06	SMPL
K3403-08	MH0CA1	4853	1.0214 -		ISM02.4	PPB	6/26/2019 15:09	SMPL
K3403-09	MH0CA2	4571	0.9585 -		ISM02.4	PPB	6/26/2019 15:11	SMPL
K3403-10	MH0CA3	7634	1.6419 -		ISM02.4	PPB	6/26/2019 15:13	SMPL
K3403-11	MH0CA4	9975	2.1642 -		ISM02.4	PPB	6/26/2019 15:16	SMPL

LB103638
INSTRUMENT ID : CV1

K3403-12	MH0CA5	4374	0.9145 -	ISM02.4	PPB	6/26/2019 15:18	SMPL
K3403-13	MH0CA6	1581	0.2914 -	ISM02.4	PPB	6/26/2019 15:20	SMPL
K3403-14	MH0CA7	485	0.0468 -	ISM02.4	PPB	6/26/2019 15:22	SMPL
K3403-15	MH0CA8	6420	1.371 -	ISM02.4	PPB	6/26/2019 15:25	SMPL
K3403-16	MH0CA9	3278	0.67 -	ISM02.4	PPB	6/26/2019 15:27	SMPL
K3403-17	MH0CB1	371	0.0214 -	ISM02.4	PPB	6/26/2019 15:29	SMPL
CCV043	CCV043	22895	5.0468 -	ISM02.4	PPB	6/26/2019 15:32	SMPL
CCB043	CCB043	72	-0.0453 -	ISM02.4	PPB	6/26/2019 15:34	SMPL
K3403-18	MH0CB3	13299	2.9058 -	ISM02.4	PPB	6/26/2019 15:36	SMPL
K3403-19	MH0CB4	2161	0.4208 -	ISM02.4	PPB	6/26/2019 15:38	SMPL
K3403-20	MH0CB5	1102	0.1845 -	ISM02.4	PPB	6/26/2019 15:41	SMPL
K3403-21	MH0CB6	13327	2.912 -	ISM02.4	PPB	6/26/2019 15:43	SMPL
K3403-22	MH0CB7	3564	0.7338 -	ISM02.4	PPB	6/26/2019 15:45	SMPL
PB120936BL	PBS936	78	-0.044 -	ISM02.4	PPB	6/26/2019 15:47	SMPL
K3404-01	MH0CB8	1708	0.3197 -	ISM02.4	PPB	6/26/2019 15:50	SMPL
K3404-02	MH0CB9	1097	0.1834 -	ISM02.4	PPB	6/26/2019 15:52	SMPL
K3404-03	MH0CC0	1835	0.348 -	ISM02.4	PPB	6/26/2019 15:54	SMPL
K3404-04	MH0CC0D	1834	0.3478 -	ISM02.4	PPB	6/26/2019 15:57	SMPL
K3404-05	MH0CC0S	14575	3.1905 -	ISM02.4	PPB	6/26/2019 15:59	SMPL
K3404-06	MH0CC1	1903	0.3632 -	ISM02.4	PPB	6/26/2019 16:01	SMPL
K3404-07	MH0CC2	2418	0.4781 -	ISM02.4	PPB	6/26/2019 16:03	SMPL
K3404-08	MH0BU7	6281	1.34 -	ISM02.4	PPB	6/26/2019 16:06	SMPL
K3404-09	MH0BZ1	3388	0.6945 -	ISM02.4	PPB	6/26/2019 16:08	SMPL
K3404-10	MH0CB2	14048	3.0729 -	ISM02.4	PPB	6/26/2019 16:10	SMPL
K3404-11	MH0CC3	27645	6.1066 -	ISM02.4	PPB	6/26/2019 16:13	SMPL
K3404-12	MH0BT9	640	0.0814 -	ISM02.4	PPB	6/26/2019 16:15	SMPL
K3404-13	MH0BU0	2754	0.5531 -	ISM02.4	PPB	6/26/2019 16:17	SMPL
K3404-14	MH0BU1	1836	0.3483 -	ISM02.4	PPB	6/26/2019 16:19	SMPL
CCV044	CCV044	23433	5.1668 -	ISM02.4	PPB	6/26/2019 16:22	SMPL
CCB044	CCB044	6	-0.06 -	ISM02.4	PPB	6/26/2019 16:24	SMPL
K3404-15	MH0BU2	1113	0.1869 -	ISM02.4	PPB	6/26/2019 16:26	SMPL
K3404-16	MH0BU3	1193	0.2048 -	ISM02.4	PPB	6/26/2019 16:29	SMPL
K3404-17	MH0BU4	876	0.1341 -	ISM02.4	PPB	6/26/2019 16:31	SMPL
K3404-18	MH0BU5	5206	1.1001 -	ISM02.4	PPB	6/26/2019 16:33	SMPL
K3404-20	MH0BV3	11840	2.5803 -	ISM02.4	PPB	6/26/2019 16:35	SMPL
K3404-21	MH0BV4	5011	1.0566 -	ISM02.4	PPB	6/26/2019 16:38	SMPL
K3404-22	MH0BV5	5927	1.261 -	ISM02.4	PPB	6/26/2019 16:40	SMPL
PB120937BL	PBS937	-9	-0.0634 -	ISM02.4	PPB	6/26/2019 16:42	SMPL
K3416-01	MH0BV6	6550	1.4 -	ISM02.4	PPB	6/26/2019 16:45	SMPL
K3416-02	MH0BV7	14465	3.1659 -	ISM02.4	PPB	6/26/2019 16:47	SMPL
K3416-03	MH0BX2	1517	0.2771 -	ISM02.4	PPB	6/26/2019 16:49	SMPL
K3416-04	MH0BX3	319	0.0098 -	ISM02.4	PPB	6/26/2019 16:51	SMPL
K3416-05	MH0BX4	423	0.033 -	ISM02.4	PPB	6/26/2019 16:54	SMPL
K3416-06	MH0BX5	388	0.0252 -	ISM02.4	PPB	6/26/2019 16:56	SMPL
K3416-07	MH0BX6	902	0.1399 -	ISM02.4	PPB	6/26/2019 16:58	SMPL
K3416-08	MH0BX7	1234	0.2139 -	ISM02.4	PPB	6/26/2019 17:01	SMPL
K3416-09	MH0BX8	877	0.1343 -	ISM02.4	PPB	6/26/2019 17:03	SMPL

LB103638

INSTRUMENT ID : CV1

K3416-10	MH0BX8D	865	0.1316 -	ISM02.4	PPB	6/26/2019 17:05	SMPL
K3416-11	MH0BX8S	14691	3.2164 -	ISM02.4	PPB	6/26/2019 17:08	SMPL
CCV045	CCV045	22962	5.0617 -	ISM02.4	PPB	6/26/2019 17:10	SMPL
CCB045	CCB045	-29	-0.0678 -	ISM02.4	PPB	6/26/2019 17:12	SMPL