



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103647

Slope : 98.2

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/26/2019	15:05
2	CAL2	1	Water	NA	NA	20.1	6.99	06/26/2019	15:06
3	CAL3	1	Water	NA	NA	20.1	10.03	06/26/2019	15:07
4	ICV	1	Water	NA	NA	20.1	7.00	06/26/2019	15:09
5	CCV1	1	Water	NA	NA	20.2	2.01	06/26/2019	15:10
6	K3541-02	1	Water	NA	NA	21.6	7.26	06/26/2019	15:11
7	K3541-03	1	Water	NA	NA	22.1	11.52	06/26/2019	15:15
8	K3541-04	1	Water	NA	NA	21.9	11.80	06/26/2019	15:16
9	K3541-04DUP	1	Water	NA	NA	21.9	11.81	06/26/2019	15:17
10	CCV2	1	Water	NA	NA	20.2	12.01	06/26/2019	15:19

LB103647

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K3541 WorkList ID : 127937 Department : Wet-Chemistry Date : 06-26-2019 14:42:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/25/2019	Water	K3541-02	Corrosivity	Cool 4 deg C	WARR01	J51	UNKNOWN-2	06/25/2019	9040C
06/25/2019	Water	K3541-03	Corrosivity	Cool 4 deg C	WARR01	J51	UNKNOWN-3A	06/25/2019	9040C
06/25/2019	Water	K3541-04	Corrosivity	Cool 4 deg C	WARR01	J51	UNKNOWN-1	06/25/2019	9040C

Date/Time 06-26-19 15:00
Received by: Jn
Relinquished by: PC

Date/Time 06-26-19 16:20
Received by: PC
Relinquished by: Jn