

AB103648

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

6/26/2019 15:48

Test: Phenolics- mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.005	0.0	0.123	
ICB1	0.005	0.0	0.015	
CCV1	1.014	0.0	0.124	
CCB1	0.006	0.0	0.015	
PB120953BL	0.003	0.0	0.014	
PB120953BS	1.018	0.0	0.125	
K3520-01	0.002	0.0	0.014	
K3520-01DUP	0.011	0.0	0.015	
K3520-01MS	0.896	0.0	0.112	
K3520-01MSD	0.874	0.0	0.109	
K3520-02	0.008	0.0	0.015	
K3520-04	-0.001	0.0	0.014	
K3520-06	0.009	0.0	0.015	
K3521-01	0.010	0.0	0.015	
CCV2	1.010	0.0	0.124	
CCB2	0.009	0.0	0.015	
K3521-02	0.011	0.0	0.015	
K3521-03	0.005	0.0	0.015	
K3521-04	0.010	0.0	0.015	
K3542-01	0.006	0.0	0.015	
K3542-02	0.012	0.0	0.015	
K3542-03	0.008	0.0	0.015	
K3542-04	0.002	0.0	0.014	
K3542-05	0.010	0.0	0.015	
CCV3	1.007	0.0	0.124	
CCB3	-0.005	0.0	0.014	

N 26
Mean 0.267
SD 0.4392
CV% 164.42

Aquakem v. 7.2AQ1

Results from time period:

Wed Jun 26 15:06:28 2019

Wed Jun 26 15:48:19 2019

AS

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0581	mg/l	6/26/2019 15:06:28	
0.1PPM	A	Phenolics-	P	0.1043	mg/l	6/26/2019 15:06:29	
0.25PPM	A	Phenolics-	P	0.2453	mg/l	6/26/2019 15:06:30	
0.50PPM	A	Phenolics-	P	0.5037	mg/l	6/26/2019 15:06:31	
1.0PPM	A	Phenolics-	P	0.9787	mg/l	6/26/2019 15:06:32	
2.0PPM	A	Phenolics-	P	2.0099	mg/l	6/26/2019 15:06:33	
ICV1	S	Phenolics-	P	1.0047	mg/l	6/26/2019 15:32:51	
ICB1	S	Phenolics-	P	0.0054	mg/l	6/26/2019 15:32:52	
CCV1	S	Phenolics-	P	1.0142	mg/l	6/26/2019 15:32:53	
CCB1	S	Phenolics-	P	0.0057	mg/l	6/26/2019 15:32:54	
PB120953BL	S	Phenolics-	P	0.0031	mg/l	6/26/2019 15:32:55	
PB120953BS	S	Phenolics-	P	1.0176	mg/l	6/26/2019 15:32:56	
K3520-01	S	Phenolics-	P	0.0016	mg/l	6/26/2019 15:32:57	
K3520-01DUP	S	Phenolics-	P	0.0108	mg/l	6/26/2019 15:32:58	
K3520-01MS	S	Phenolics-	P	0.8965	mg/l	6/26/2019 15:32:59	
K3520-01MSD	S	Phenolics-	P	0.8737	mg/l	6/26/2019 15:33:00	
K3520-02	S	Phenolics-	P	0.0077	mg/l	6/26/2019 15:33:01	
K3520-04	S	Phenolics-	P	-0.001	mg/l	6/26/2019 15:33:02	
K3520-06	S	Phenolics-	P	0.0092	mg/l	6/26/2019 15:42:53	
K3521-01	S	Phenolics-	P	0.0095	mg/l	6/26/2019 15:42:54	
CCV2	S	Phenolics-	P	1.0104	mg/l	6/26/2019 15:42:55	
CCB2	S	Phenolics-	P	0.0086	mg/l	6/26/2019 15:42:56	
K3521-02	S	Phenolics-	P	0.0107	mg/l	6/26/2019 15:42:57	
K3521-03	S	Phenolics-	P	0.0047	mg/l	6/26/2019 15:42:58	
K3521-04	S	Phenolics-	P	0.0101	mg/l	6/26/2019 15:42:59	
K3542-01	S	Phenolics-	P	0.0063	mg/l	6/26/2019 15:43:00	
K3542-02	S	Phenolics-	P	0.0122	mg/l	6/26/2019 15:43:01	
K3542-03	S	Phenolics-	P	0.008	mg/l	6/26/2019 15:43:02	
K3542-04	S	Phenolics-	P	0.0019	mg/l	6/26/2019 15:43:03	
K3542-05	S	Phenolics-	P	0.0104	mg/l	6/26/2019 15:43:04	
CCV3	S	Phenolics-	P	1.0068	mg/l	6/26/2019 15:48:17	
CCB3	S	Phenolics-	P	-0.0045	mg/l	6/26/2019 15:48:18	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

6/26/2019 15:07


Test Phenolics-

Accepted 6/26/2019 15:07

Factor 9.187
Bias 0.014

Coeff. of det. 0.999759

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.020	0.0581	0.0500	
2	Phenol-2	0.025	0.1043	0.1000	
3	Phenol-2	0.041	0.2453	0.2500	
4	Phenol-2	0.069	0.5037	0.5000	
5	Phenol-2	0.121	0.9787	1.0000	
6	Phenol-2	0.233	2.0099	2.0000	